

R-Log



User's manual

Update 12/19/2013

Copyright 2011-2013 LSI LASTEM. All rights reserved.

This manual can be modified without notice.

Anybody can copy, print or publish this manual without LSI LASTEM written authorization.

LSI LASTEM reserves the right to modify the product without an immediate revision of this document.

Sommario

1.	General safety rules.....	5
2.	Foreword	6
2.1.	Disposal.....	6
2.2.	How to contact LSI LASTEM	6
3.	Guide to the start.....	7
3.1.	Description	7
3.2.	Models.....	8
3.3.	Mechanical and electrical installation.....	9
3.3.1.	Instrument power supply.....	9
3.3.2.	Internal rechargeable battery.....	10
3.3.3.	Inputs and actuators	10
3.3.4.	Single-ended inputs.....	12
3.3.5.	Serial communication lines	13
3.3.6.	Modem power supply.....	13
3.4.	Configuring the operative mode	14
3.4.1.	Language configuration	14
3.5.	Multi-point system	14
4.	Instrument use.....	16
4.1.	Frontal panel overview.....	16
4.2.	Using the keyboard	16
4.3.	Display informations and controls	18
4.3.1.	Input of commands with immediate effect	18
4.3.1.1.	Power on and power off the instrument.....	18
4.3.1.2.	Fast acquisition mode.....	18
4.3.1.3.	Configuration with automatic recognition of probes	19
4.3.1.4.	Record the starting and ending time of data acquisition (Survey n).....	19
4.3.1.5.	Reset statistics	19
4.3.1.6.	Fast transmission mode.....	20
4.3.1.7.	System clock setting.....	20
4.3.2.	Product presentation window.....	20
4.3.3.	Instantaneous values of measures	21
4.3.4.	Diagnostic information	22
4.3.5.	Shutting off the display	28
4.4.	Reset button.....	28
4.5.	LED's.....	29
5.	R-Log additional features.....	30
5.1.	Starting and running the survey	30
5.1.1.	Automatic recognition of probes.....	30
5.1.2.	Checking the power supply voltage	32
5.2.	Acquisition and calculation of measures	32
5.2.1.	Acquisition from sensors with serial or radio output.....	33
5.2.2.	Acquisition from thermocouples.....	34
5.2.3.	Details about the measure acquisition process.....	34
5.2.4.	Acquisition from status signals	34
5.2.5.	Fast acquisition mode of the measures	35
5.2.6.	Activation of sensor control.....	35
5.2.7.	Details about calculated measures	36

5.3.	Measure elaboration	36
5.3.1.	Vectorial calculation specific for anemometric quantities.....	37
5.4.	Storing elaborated data.....	38
5.4.1.	Memory autonomy	38
5.5.	Actuation logics	39
5.5.1.	Eolic alarm	40
5.5.2.	Evaporimeter filling	40
5.5.3.	Start precipitation alarm.....	40
5.5.4.	Flood alarm	41
5.5.5.	Threshold value compare	41
5.5.6.	Timer.....	44
5.5.7.	Snow level alarm.....	45
5.5.8.	System error	45
5.6.	Communication modes	46
5.6.1.	RS232 serial port.....	46
5.6.2.	ZigBee Radio	47
5.6.2.1.	Network set-up	49
5.6.3.	Ethernet	51
5.6.3.1.	DEA550 converter configuration.....	51
5.6.3.2.	XPORT configuration (internal Ethernet port)	51
5.6.4.	Data trasmission through GPRS connection	54
5.6.4.1.	Use of DEA717 GPRS Industrial Plus modem.....	56
5.6.4.2.	Use of DEA718 GT863-PY Telit modem.....	56
5.7.	Operating of low energetic consumption	57
6.	Appendixes.....	58
6.1.	Technical specifications	58
6.2.	Function library for derived calculations	61
6.3.	Error messages	63
6.3.1.	Disabling error indication	64
6.3.2.	Error found in measure	64
6.3.3.	Low battery alarm	65
6.4.	Instrument maintenance	65
6.5.	Declaration of conformity	66

1. General safety rules

Please read the following general safety rules in order to avoid injuries to people and to prevent damages to the product or to products that may be used in connection with it. In order to avoid damages, use this product exclusively according to the instructions herein contained.

Installation and maintenance interventions are to be exclusively carried out by authorized and skilled people only.

Install the instrument in a clean, dry and safe place. Humidity, dust and extreme temperatures may deteriorate or damage the instrument. In such cases, we advise installing the instrument inside a suitable container.

Power the instrument in a suitable manner. Connect the instrument to the power supply indicated in the model in your possession.

Carry out all connections in a suitable manner. Pay strict attention to the connection diagrams supplied with the instrument.

Do not use the product in case of suspected malfunction. In case of suspected malfunction, do not power the instrument; contact authorized technical support immediately.

Before every maintenance of electrical connections, power supply, sensors and computer-equipments:

- **disconnect the power supply,**
- **discharge the electrostatic discharges touching one conductor or one earth apparatus.**

Do not use the product in the presence of water or condensing humidity.

Do not use the product in a potentially explosive atmosphere.

Internal lithium battery. Do not replace the battery with wrong type. Possible explosion risk.

2. Foreword

R-Log datalogger has been designed for environmental applications; it comes from E-Log primarily and utilizes the same use modes and management programs (3DOM, Gidas ...).

Thanks to its low power consumption, its sizes, the range of signals that it can receive, the protection as regards heavy environmental conditions and possible overvoltages, R-Log datalogger is particularly suitable for measurements of weather, water, air quality, indoor and outdoor environmental monitoring applications. Besides, as R-Log is an instrument equipped with built-in radio system that runs according to ZigBee (IEEE 802.15.4) international standard, it's particularly suitable to configure monitoring networks and support the implementation of master-slave configurations; so the radio grants to our datalogger, in addition to well-known *multi-measurement-system* characteristic also the characteristic of *multi-point-system*.

R-Log line is divided in two big typologies (you can find the detailed list of models in §3.2):


- R-Log with inputs suitable to receive sensors with connectors for automatic configuration by means of recognition resistance; it's suitable for indoor applications (microclimate).
- R-Log with terminal board inputs; it's suitable for outdoor applications (meteorology).

R-Log can be fitted with a wide range of accessories to enhance its power autonomy, to protect it against extreme weather conditions and for data transmission via RS232/485, USB, Ethernet, modem PSTN/GSM/GPRS.

2.1. Disposal

R-Log is a highly electronic scientific device. In accordance with the standards of environmental protection and collection, LSI LASTEM advises to handle R-Log as waste of electrical and electronic equipment (WEEE). It is therefore not to be collected with any other kind of waste.

LSI LASTEM is liable for the compliance of the production, sales and disposal lines of R-Log, safeguarding the rights of the consumer. Unauthorized disposal will be punished by the law. Dispose of the dead batteries according to the regulations in force



2.2. How to contact LSI LASTEM

In case of problems contact the LSI LASTEM technical support at support@lsi-lastem.it, or fill in the *On-line technical support request* form accessible from the home page of the website www.lsi-lastem.it.

For further information:

- | | |
|-----------------------|---|
| • Telephone | +39 02 95.414.1 |
| • Address | Via ex S.P. 161 – Dosso n. 9 - 20090 Settala Premenugo, Milano |
| • Home page | www.lsi-lastem.it |
| • Sales | info@lsi-lastem.it |
| • After-sales service | support@lsi-lastem.it , riparazioni@lsi-lastem.it |

3. Guide to the start

3.1. Description




R-Log is leading-edge datalogger worked out by LSI LASTEM, and thanks to its multi-measurement and multi-point system it's suitable for all environmental applications.


The front panel of instrument is equipped with illuminated display (4x20 characters), signalling bright leds, function and arrow keys.

On top side of instrument there are the serial communication port (equipped with a double connector for applications where you want to interface R-Log with a DCE device or a DTE device), the Reset key, the socket for connection of external power supply system and the aerial of radio that's mounted inside instrument. When you unpack the instrument make attention that aerial could has been bent (owing to transport); for optimal operation of datalogger put aerial in vertical position handling it with care without any pressures or sharp movements.

On bottom side we can find the inputs for connection of sensors; according to R-Log model, for connection of sensors, it's possible to find fix mini-din connectors (models series ELR510), or usual terminal board with pluggable terminals (models series ELR515); in last case extract the terminals from their housing using a suitable tool.






Terminal board with fix mini-din connectors (models series ELR510)



Terminal board with pluggable terminals (models series ELR515)

On back panel of instrument there is a label that carries, in addition to standard information about power supply and radio features, also following information:

- R-Log model;
- Serial number of instrument (univocal number consisting of 8 numerics that marks each datalogger);
- MAC address for models ELR516M, or equipped with Ethernet output;
- Free space for annotation of used configurations and their connections with sensors connected to datalogger.

R-Log Radio data logger									
<input type="checkbox"/> ELR510M	<input type="checkbox"/> ELR515M	12V --- /600mA							
<input type="checkbox"/> ELR510R	<input type="checkbox"/> ELR515R	Radio freq: ISM 2.4GHz; Tx output power: 10mW							
<input type="checkbox"/> ELR510S	<input type="checkbox"/> ELR515S	MAC address :							
<input type="checkbox"/> ELR516M		0	0	2	0	4	A	A	9
Cfg. name	Sonde/Probes								
	IN1	IN2	IN3	IN4	IN5				
  									
LSI-LASTEM S.r.l. - Via ex S.P.161 Dosso 9 - 20090 Settala (MI) - Italy www.lsi-lastem.it									

3.2. Models

All the R-Log models are equipped with alphanumeric display, keyboard and inner battery; their distinguishing properties are as follows:

- Presence of connector inputs or terminal board inputs;
- Operating mode: Master, Repeater, Slave (see §5.6.2 for features of different operating mode).

Code	Power supply 12 Vdc	Connector inputs	Terminal board inputs	MASTER mode	REPEATER mode	SLAVE mode	Ethernet output
ELR510M	X	X	-	X	-	-	-
ELR510R	X	X	-	-	X	-	-
ELR510S	X	X	-	-	-	X	-
ELR515M	X	-	X	X	-	-	-
ELR516M	X	-	X	X	-	-	X
ELR515R	X	-	X	-	X	-	-
ELR515S	X	-	X	-	-	X	-

- **ELR510M**: connector inputs, MASTER;
- **ELR510R**: connector inputs, REPEATER;
- **ELR510S**: connector inputs, SLAVE;
- **ELR515M**: terminal board inputs, MASTER;
- **ELR516M**: terminal board inputs, MASTER, Ethernet output;
- **ELR515R**: terminal board inputs, REPEATER;
- **ELR515S**: terminal board inputs, SLAVE.

3.3. Mechanical and electrical installation

R-Log is able to be used for both internal use (placed on a flat surface or fixed onto a wall) and external use (inside suitable protection boxes).

The use of supports furnished by LSI LASTEM is advisable for indoor applications:

BVA304	Universal tripod to support BVA311 stand
BVA311	Stand for support of n.5 probes and R-Log on tripod or on wall
BVA312	Stand for support of further n.5 probes and R-Log on BVA311 or on wall
BVA314	Sidebar for support of stand on the wall

3.3.1. Instrument power supply

R-Log is equipped with 1,95 A, lithium ion, internal rechargeable battery.

All models can be powered by an external power supply pack (8 ÷ 14 Vdc), by means of connector placed on top panel of instrument; in this case the positive pole is pole inside the connector. In any case pay attention not to invert the power supply polarity (although the instrument is protected from wrong procedure).

For ELR515 and ELR516 models is also possible to power by means of terminal board.

Make reference to table below for connection with instrument-input and sensors-output power-supply terminals or other devices that need power supply.

Connection	Terminal
0 Vdc battery	28 or 29
+ Vdc battery / + Vdc fixed to power sensors and external devices	30 or 31
GND	32

When there's GND wire (grounding), it's recommended to connect it with 32 terminal.

In case of no switch on of the device, push the RESET button (located on the higher panel of the instrument) and then try to push the keys combination for switch on; if the sensor is still off, send it for repair at LSI LASTEM.

WARNING: in case terminals 5-6 and 16-17 feed outside equipments, they must be equipped with power fail circuit against short circuits or absorbed currents above 1 A.

WARNING: in applications where there is LSI LASTEM *BSO103.1* sensor (CO₂ probe), R-Log datalogger **must** be powered by means of external power supply pack; the only R-Log internal battery isn't able to supply enough power for right operation of connected sensor.

3.3.2. Internal rechargeable battery

The internal rechargeable battery ensures to the instrument the autonomy, the independence and the handiness during different applications that involve it; it is normally subject to unload. The low battery alarm occurs on datalogger through three times blinking of LED ■-Err (see §6.3.3).

The total discharge of the battery will lose the date/time of the datalogger; this condition is very dangerous because it makes the stored data unreadable so also compromising the operation of data unloading and data storing on the acquisition system.

Note: It's advisable to check the date/time displayed by the datalogger and eventually update to the current date/time through 3DOM software (see §4.5.2 of SWUM_00339 manual reported on LSI LASTEM products DVD MW6501):

- before starting the survey;
- before and after operations of battery recharge;
- after a long period of inactivity;
- when the message of low battery is displayed (see §6.3.3).

3.3.3. Inputs and actuators

The instrument is fitted with 4 analogue inputs in *differential* mode and 8 in *single-ended* mode (§3.3.4 **Errore. L'origine riferimento non è stata trovata.**) (independent, for sampling of voltage signals, current, resistance, Pt100 with three or four wires, thermocouples), a digital input (for sampling of pulse inputs, frequency or digital state) and 2 actuators used to power the sensors connected to the terminal board; the actuators can also be used by the actuation programmable logics, than can produce alarms according to the values acquired by the sensors . The voltage available on these terminals depends on the kind of power supply received by the instrument.

Use the program *3DOM* (see SWUM_00339 available on LSI LASTEM product DVD – MW6501) to configure the operation to switch on the sensors by means of the switching powers. In order to choose the actuation time is advisable to consider both the energetic saving and the time the sensor needs to initialize.

The association between input and switching power is fixed, as shown in the table below. The number of the terminal is indicated in *italics*; we understand, for example, that inputs 1 and 2 both make use of the first actuator; therefore it cannot be used for the other inputs. In case of sensors that generate two signals (like the thermo-hygrometric sensor), it's suitable to select both inputs that use the same actuator.

A third actuator is on the pin9 of the serial port connector and it can be used for sensors that interface to R-Log with serial connection; if configured appropriately with a logic based on the power level or battery power level, it can be used as fixed power supply.

All the actuators can be activated manually by the diagnostic window type 4 (see §4.3.4).

The following tables show the terminals used to connect sensors and power supply for ELR515 and ELR516 models.


TERMINAL BOARD								
Analogue input	Signal				GND	Probe switched power / Actuator		
	A	B	C	D		Number	+V	0 V
1	1	2	3	4	7	1	5	6
2	17	18	19	20				
3	8	9	10	11	23	2	21	22
4	24	25	26	27				

Digital input	Signal			GND
	E	F	G	
5	12	13	14	15 or 16

Actuator/Output alarm on the serial	
+V	0V
Pin 9	Pin 5

The following pictures explain in detail the connections of all types of sensors, both analogue and digital.


Sensors with analogue signal (in differential mode):



The drop resistance, indicated by Rx, is used to return a voltage signal from the current generated by the sensor. Program 3DOM supplies a library for the setting of the LSI LASTEM sensors,

including some powered outputs; for such models the settings have been arranged to use the energized scale $-300 \div 1200$ mV, thus being able to use 50Ω drop resistances.

Sensors with digital signal:



3.3.4. Single-ended inputs

Starting from datalogger firmware version 2.10.00 it is possible to double the number of analogue inputs, passing from 4 to 8. This function is available only for sensors with tension or current signals (this function is called single-ended). Resistive signals, instead, go on occupying a full single physical input (this function is called differential).


Differential inputs are more immune from electromagnetic disturbances than single-ended ones and, when possible, they have to be preferred.

All sensors generating current signals and all powered sensors can be connected to single-ended inputs (see connection scheme).

WARNING:

To R/M Log instruments it is possible to connect only: radiometers, thermo-couples, fluxmeters or signals externally powered (see connection scheme)

Sensors with analogue signal (single-ended mode):



The drop resistance, indicated by Rx, is used to return a voltage signal from the current generated by the sensor. Program 3DOM supplies a library for the setting of the LSI LASTEM sensors, including some powered outputs; for such models the settings have been arranged to use the energized scale $-300 \div 1200$ mV, thus being able to use 50Ω drop resistances.

When a probe is added using the 3DOM probes library, those sensors are always added in differential mode; for this reason is not possible add a single-ended probe to a configuration where there is only one single-ended sensor free; when this situation occurs set the measure parameters manually matching the single-ended probe needs; otherwise change, if possible, different measures from differential to single-ended mode until obtaining a complete differential input free (two single-ended sequential inputs).

3.3.5. Serial communication lines

The instrument is fitted with one serial port; there is another serial port but it is dedicated exclusively to the internal radio and not visible outside. The available serial port, located on the upper panel of the instrument and with a double connector (female if the datalogger is used as a DCE device, male if it is used as a DTE device), is used to program the instrument's operative modes and data download, by means of the LSI CISS communication protocol.

The serial port leaves our factory with the following default configuration:

- Baud rate: 57600 bps;
- Data bit: 8;
- Stop bit: 1;
- Parity: none;
- Network address: 1.

The instrument can be programmed; however only the *baud rate* and *network address* can be changed; all other parameters cannot. The rate can be programmed from 1200 to 115200 bps; in case of use of phone modem GSM/GPRS it is suitable program 9600 bps on bound rate.

Electrically speaking, both ports are configured as a DCE device. The following table shows the meaning of each serial connectors' pin:

RS232 Serial port	Signal	Pin for DCE device	Pin for DTE device
	TD	2	3
	RD	3	2
	GND	5	5
	RTS	8	7
	CTS	7	8

3.3.6. Modem power supply

R-Log can power the modem continuatively (through 12 Vdc powering from the terminals, like shown at §3.3.3), or through timed actuator, in order to reduce the electric energy consumption of the system.

You must use actuator n.3 to power the modem through actuator. Activate it according to the different connected modem:

- GSM Modem: the actuator is started at time of instrument starting; in this case the telephone connection holds on also during the remote re-configuration of the instrument through 3DOM program; the next shutdown happens by means of the timed actuation logic,

according to the programmed shutdown time; i.e. in case of following programmed timed logic, starting at 15:00 o'clock and shutdown at 16:00 o'clock, if the instrument is started at 14:30, the effector will operate one hour and half;

- GPRS Modem: the actuator is started at time of instrument starting and the shutdown happens about one minute later; next modem activation will happens according to the GPRS data transmission timing, as programmed with *3DOM*.

3.4. Configuring the operative mode

The instrument supplied by LSI LASTEM is equipped with standard configuration that includes only the battery tension and the internal temperature; for models series ELR510 it's possible to set-up the configuration automatically, as they are equipped with function for sensors auto-recognition; but configuration of models ELR515 and ELR516 must be set-up by means of *3DOM* software, through the direct input from software library or through support of suitable models designed for different user applications (microclimate, walls transmittance ...).

Both the measurements autogeneration and the applicative models produce the generation of only type "Ave" elaborations (mean value). When configuration is set-up starting from sensors in *3DOM* library, it's possible to obtain the generation of more complete processings ("Min/Ave/Max", "Tot", "Stdev", "%valid data" ...); so in this case it'll be possible to delete, manually, the elaboration you aren't interested in.

For both R-Log releases the modification of configuration, in order to obtain new setup satisfying with one's own needs, is made through PC connected to serial communication port with *3DOM* program. For details about all settable parameters, make reference to online help of this program or user manual DVDs made by LSI LASTEM. For connection of serial line and its programming, see §3.3.5.

During the configuration data upload (by means of the *3DOM* program) the survey (in progress in the instrument) is closed, in order to allow to the instrument the right reconfiguration condition. At this point the measures list on the display (see §4.3.3) is replaced with the notice "Survey halted"; it means that the survey has been halted.

Note: when programming a new configuration, all data stored in the instrument's memory will be deleted. For this reason we advise to receive on PC all data from the instrument before uploading the new configuration.

3.4.1. Language configuration

The instrument always uses English language for display messaging; for measurement texts it is factory programmed to use Italian names but this can be changed by *3DOM* configuration program.

3.5. Multi-point system

It's defined:

- *NETWORK*, the coverage of an area through radio frequency devices, that can receive, transfer, broadcast messages (in repeater bridge mode) from and to other points of the same network;
- *NODE*, every active element the network consists of .

A lot of environmental applications need the possibility to set-up a system that can monitor several points contemporaneously for a given period of time. Using ZigBee technology, R-Log represents the *node* of one *network* perfectly, and in this way it allows to the user to set-up a multi-point system in an easy way, in other words a wide monitoring network that can keep several measurement positions under control.


Placing a certain number of R-Log accurately, determined according to typology of environment that has to be monitored, it's possible to obtain one net that satisfies the monitoring characteristics required by the application.

For more details about ZigBee network characteristics and for suggestions about setup of one network, see §5.6.2.

4. Instrument use

4.1. Frontal panel overview

Picture below shows the frontal panel of the instrument:



The top side of frontal panel houses the bright indicators (power supply, battery, transmission and reception, error signalling) and the display, that can be used for right operation of instrument and for display of data acquired by sensors in real time.








The bottom side of frontal panel houses the keyboard that includes function keys and arrow keys.

4.2. Using the keyboard










The keyboard includes a series of four arrow keys and four function keys.

The following table summarizes the main functions of each key, contextually according to the state the instrument finds itself in.

While showing the values of the measures:

	Goes to the diagnostic window Type 1
	Changes the display mode of the measure name (extended, abbreviated), of the measure unit, the number of terminal which the sensor is physically connected to or ID/channel of radio sensor, alarms set and resistances of recognition (only for series models ELR510)
	
	Scrolls up the measure list
	Scrolls down the measure list
	Holds/releases measure scrolling
	Switches the display off and on

While showing diagnostic information:


	Shifts to the display window of the values of the measures.
	If applicable, pressed at the same time with  key, it resets display data (statistics, errors and other information).
	In the display window of the communication statistics it shifts from serial 1 statistics to serial 2 and vice versa; in the type 5 diagnostic window it shows or hides the additional information about the GPRS modem (only LSI LASTEM technicians).
	
	Shows the previous diagnostic window.
	Shows the next diagnostic window.
	In the type 5 diagnostic window it switches on the GPRS modem (in case it's switched off) and starts the data communication through GPRS; in the type 4 diagnostic window it powers on actuators; in the type 7 diagnostic window it powers on the radio of Slave instruments.
	In the type 5 diagnostic window it resets GPRS modem.

4.3. Display informations and controls



Through the user interface of the instrument, is possible:




- Input of commands with immediate effect
- Display of:
 - Product presentation window;
 - Instantaneous values' scrolling list of all programmed measures;
 - Diagnostic information.

4.3.1. Input of commands with immediate effect

R-Log  key gives the double function to key pressed with it contemporaneously.



4.3.1.1. Power on and power off the instrument


When the instrument is off, pressing  and  keys contemporaneously, the instrument turns on and shows the product presentation mask (see §4.3.2) that will be displayed for about 10 seconds; when this mask disappears, the datalogger will start the survey according to the configuration saved before the shutdown.

When the instrument is on, pressing the same ( and ) keys, the instruments shows the mask below and will be awaiting for confirmation through  key before its definitive shutdown; at pressing of any other key the instrument will return to previous condition.

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
P	r	e	s	s		F	n	t	o		c	o	n	f	i	r	m		



4.3.1.2. Fast acquisition mode

Pressing  and  keys contemporaneously from instantaneous data display mask, it's possible make sensors acquisition rate faster than rate set-up by configuration.

Pressing again  and  keys contemporaneously the acquisition rate will return to follow the configuration setup..

For more details see §5.2.5

4.3.1.3. Configuration with automatic recognition of probes

After physical connection of sensors, it's possible to configure the instrument (only for R-Log ELR510 model) with automatic recognition of probes by pressing  and  key contemporaneously; R-Log will show the following mask:

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
P	r	e	s	s		E	n	t	e	r		t	o						
c	o	n	f	i	r	m		a	u	t		o	-	c	o	n	f	i	g
W	A	R	N	!		S	t	o	r	e	d			e	l	a	b	.	
w	i	l	l		b	e		d	e	l	e	t	e	d	!				

and stay awaiting for confirmation through  key before you start the operations for auto-recognition of probes.



At the end of auto-configuration function, if the instrument hasn't found any errors, a new survey will be started and new mask of instantaneous values will be displayed.

Warning: this operation will delete all stored data; before you confirm the auto-recognition, download the stored data on PC.



4.3.1.4. Record the starting and ending time of data acquisition (Survey n)

(This function is valid only for R-Log configured with MASTER modes)



R-Log is a mono-survey instrument, but offers to the user the possibility to mark the time of surveys; in other words it allows to fix manually, for one survey, the starting and ending time of data acquisition. These information are shown by LSI LASTEM application programs (such as Microclima) in order to make easy the selection of data included into given period of sampling.

The starting time, to consider data, is fixed by pressing  and  keys at the same time; the display will show the following mask:





1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
			-	-		S	u	r	v	e	y		n		-	-			
1	T	e	m	p															
2	V	B	A	T	T														
3	e	c	c	.															

But pressing again  and  keys at the same time it'll be fixed the ending time; the instrument will continue to save the data, but not inside survey closed before. An univocal numeric value n will identify the selected slot; n will be increased every time a new slot for acquired data consideration is fixed. The reset of survey number is made only modifying configuration from PC or starting new auto-recognition procedure.

4.3.1.5. Reset statistics

Pressing  and  keys at the same time, it's possible to reset values of diagnostic mask statistics number 2 (see §4.3.4), in other words the values referred to communication (bytes and transferred and received messages).

4.3.1.6. Fast transmission mode







Pressing  and  keys at the same time from diagnostic window type 2, it's possible select the sensor fast transmission mode (transmission every second). This function will be active **only** if a value different from 0 is set in the *Serial communication port x* → *Instantaneous values automatic transmission rate* option (x takes values 1 or 2 according to the relative serial port) in the configuration send to the instrument. The fast transmission mode will be activate only on the serial port that is interrogated during the display of diagnostic window: when you display the diagnostic window type 2 relative to the serial port 1 (C1), the keys combination  and  will activate the fast transmission mode only for communication serial port 1.

Pressing again the same keys combination, the transmission mode will return to work according the configuration's parameters.

4.3.1.7. System clock setting

A few seconds after being switched on, the instrument can require to re-set the clock (date and time). This can happen only in particular situations as, for example, the complete discharge of the battery: in this case the date/time setting is compulsory and cannot be cancelled.

The date/time changing can anyway be performed whenever desired from the Diagnostic window type 1. Use the following keys:

- : enter *change* mode; exit from *change* mode confirming and storing the new date/time;
- : exit from *change* mode discharging possible modified values;
-  : move the cursor over the field to modify;
- : it augments the value selected by the cursor; other elements of date/time can change contemporarily;
- : it reduces the value selected by the cursor; other elements of date/time can change contemporarily.

4.3.2. Product presentation window

When the instrument is started, the following informations will appear on the display for a couple of seconds:

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
L	S	I		L	A	S	T	E	M						I	t	a	i	y
R	-	L	o	g	x	x	x	z			V	a	a	.	b	b	.	c	c
S	N		y	y	m	m	n	n	n	n	/	u	u	u	u	u	u	u	u

where:

- xxx: instrument model;
- z: radio ZigBee model (m Master, r Repeater, s Slave);

- aa.bb.cc: program release (higher.lower.build);
- yymnnnnn: serial number;
- uuuuuuuu: serial number or instrument number (settled by the user).

These informations are also available while the instrument is working and can be called by simply using the keyboard to select this window to be displayed.


4.3.3. Instantaneous values of measures

Each line displays the last value of a measure, be it acquired or calculated. The information is displayed as follows (one or more lines):

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
n	n	m	m	m	m	m	m	m	m	m	m	m	v	v	v	v	v	v	v

where:


- nn: ordinal number of the measure; it doesn't indicate the physical input number;
- mmmm...: completed measure name; see §3.4.1 to change this text or the language used;
- vvvvvvv: value of measure; the error state is identified by the writing "Err"; the value is justified on the right.


Use navigation key  to display the abbreviated name of the measure, the measured value and its measure unit.

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
n	n	m	m	m	m	m	z	z	z	z	z	z	z	v	v	v	v	v	v

where:

- nn: ordinal number of the measure; it doesn't indicate the physical input number;
- mmmmm: shortened measure name (5 characters);
- zzzzzzz: value of measure; the error state is identified by the writing "Err"; the value is justified on the right;
- vvvvvv: measure unit.

Pressing again  key you obtain the display with shortened name of measured quantity and number of input the sensors is connected; if the measurement has been calculated, instead of number of input it's displayed sign "-"; if measurement has been acquired by serial port or radio, instead of number of input they're shown Network Address of sensor (with prefix "A"), and number of its channel to which the measurement is referred (with prefix "C").


In case the instrument has been programmed with one actuation logic at least, a further pressing of  key displays the alarm condition of measurement: sign "OK" shows that measurement hasn't generated any alarm condition, at the opposite it's displayed sign "Alarm".

In this mask, in R-Log models of series ELR510, it's also displayed value of resistance that has allowed the operation of auto-recognition of probe through datalogger.

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
n	n	m	m	m	m	m	z	z	z	z	u	u	R	R	v	v	v	v	v

where:


- nn: ordinal number of measurement; it doesn't specify number of physical input;
- mmmmm: shortened name of measurement (5 letters);
- zzzz: values that specify if measurement has been acquired directly (it'll be displayed letter "T"), calculated (it'll be displayed symbol "-"), acquired through radio (it'll be displayed letter "A" plus address and letter "C" plus number of channel to which it's associated);
- uu: value of physical input to which probe is connected;
- vvvvvv: value of recognition resistance.



keys allow to shift from one format to an other one.

4.3.4. Diagnostic information

Some diagnostic windows are used to display information about the operation of the instrument and


the statistics; they are displayed by the pressing the  button from the window of *Instantaneous values of measure*. Diagnostic window type 1:


0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
D	T		d	d	/	m	m	/	y	y		h	h	:	m	m	:	s	s
E	r	r		e	e	e	e	e	e	e	e								
M	e	m		m	m	m	m	k	B		F	r	e	e	w	w	w	%	
M	s	r		A	a	a		S	s	s		C	c	c					

where:

- dd/mm/yy hh:mm:ss: current date/time of the system;
- eeee...: numeric code, expressed in hexadecimal number, which corresponds to the 32 bits of the error window of the system; in order to decode this error, go to §6.3;
- mmmm...: data memory capacity in kByte;
- www: percentage of available memory; R-Log has circular storage and for this reason the memory is virtually end-less. The calculation of availability is made according to total memory space dedicated to elaborated data in comparison with data already transferred to PC from serial line; during the reading of elaborated data this value is updated continuously in order to specify the percentage of available memory in relation to datum is being read from one instant to the next; if the instrument shows that percentage value is zero, it means that circular storage algorithm has stored new data and deleted the older ones.
- aa: number of measures acquired by the inputs of the instrument;
- ss: number of acquired measures from serial port;
- cc: number of calculated measures.






By pressing the arrow  you can see the old camp which is the number in hexadecimal notation that expresses the value of the position in writing in the memory of the elaborated data (with a starting value which is equal to 0); when using a 2 MB memory, this value is expanded to 1FFF; the unitary increment of this value indicates a 256 byte consumption.

Use key  to shift to diagnostic window type 2:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
R	e	s	e	t		d	/	m	m	/	y	y		h	h	:	m	m	
C	n		B	y	t	e						M	s	g					
R	x		a	a	a	a	a	a	a			b	b	b	b	b	b	b	b
T	x		c	c	c	c	c	c	c			d	d	d	d	d	d	d	d

where:

- dd/mm/yy hh:mm: date/time the statistic was last reset; resetting the statistical values can take place locally, (by pressing at the same time of  and  keys) or remotely through a suitable PC command;
- n: number of the instrument's serial port; go to the following line with  key;
- aaaaaaaaa: number of bytes received;
- bbbbbbbb: number of messages received;
- cccccccc: number of bytes transmitted;
- dddddddd: number of messages transmitted;

This mask is very important for diagnostic reasons, in order to value the quality of communication among R-Log datalogger and between R-Log and PC.

Communication between PC and R-Log


Every standard message consists of about 13 characters; so:

In case of reception of more than 13 Rx characters and no increase of Rx of Msg value	R-Log bit rate could be higher than rate set on PC
In case of reception of less than 13 Rx characters and no increase of Rx of Msg value	R-Log bit rate could be lower than rate set on PC
In case of rates set correctly, right reception of characters but not decoded message (Msg value doesn't increase)	Could have been used different protocols or could be communication electric problems (i.e. on 485 line, + and - could have been inverted) or communication line is too jammed.
In case of: rates set correctly, right reception of characters, encoded message but no answer (Tx of Msg value doesn't increase)	The addressee of message isn't stated R-Log but an other datalogger: check network address (Protocol ID)

Communication among R-Log dataloggers

Remember that radio communications among R-Log dataloggers go only through Serial 2. So observing the values of this diagnostic mask you could see the increase of both Tx and Rx bytes values, as radio and processor of same R-Log communicate together; on the contrary Msg values increase in different way according to type of datalogger (master or slave), as the communication from slave to master is mono-directional (slave instrument sends messages to master in automatic way and the latter doesn't send any answer to sender). So:

- in an R-Log master will be increased only Rx value of Msg field;
- in an R-Log slave will be increased only Rx value of Tx of Msg field.



Use  key to shift to diagnostic window type 3:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
C	1		A	d	d	r	x	x	x	>	y	y	y	y	y	y	b	p	s
C	2		A	d	d	r	x	x	x		y	y	y	y	y	y	b	p	s

where:



- xxx: network address of the instrument;
- yyyy: communication rate (bit rate) of the serial port.

The character “>” shows the current serial port; for this port are valid the modification commands


of the bit rate (from 1200 to 115200 bps), using the keys   on instrument; this modify is possible only for serial port number 1.

The modify of the rate of communication serial port number 2 (that exclusively used by radio) is possible only by 3DOM software.

R-Log serial port are standard set to 115200 bps.

By pressing key  and  the instrument will set the network address temporarily to value 1 and will carry out, after some seconds, at least one test transmission of the instantaneous values that can be checked by means of any program for terminal emulation for diagnostic purposes.




Note: the communication values can be changed by means of keyboard but this is a transitory modify, because it's arranged to solve quickly possible communication problems with the outside equipments; in fact every time the instrument is switched on again it uses the PC's programmed configurations. Use 3DOM program to program these setups definitively.

Use  key to shift to diagnostic window type 4:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
O	U	T		1	2	3													
				x	x	x													
				^															
B	a	t	t.	:		y	.	y	V	(x	x	x		%)			

where:

- x: value of the single actuation output: 0 = output is inactive, 1 = output is active.
- y.y: value of the power supply-voltage measured by the instrument (Volt)
- ^ indicates the selected actuator;
- xxx: percentage value of power battery of the instrument. When the datalogger is externally powered displays a value below 100% (60% typically) even when the battery is fully charged (can be determined by switching off the charging LED on the frontal panel); removing the external power supply the value of battery is shown correctly. This feature is similar when “battery voltage” is a sampling measure.




For diagnostic uses, use the   keys to shift on the selected actuator (indicated by ^ symbol), and  key to shift from switched on to switched off modes and viceversa .


From 2.3 firmware version it is possible to force the actuator state among to the internal logics that could change (actuators logics or sensor power supply logics); in addition to the 0 value (actuator off) and 1 value (actuator on) is shown the F letter, one for each actuator, if the forcing is active. In practice, if the forcing is also imposed to an actuator state (that can assume 0 or 1 value) means that actuator will never be change by any internal logic (except the commands received from remote through CISS communication protocol or Modbus that have priority on F status); if the F forcing is off, the internal logic can normally change the actuator status.

The diagnostic window is so displayed:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
						f	f	f											
						x	x	x											
O	u	t				1	2	3											
B	a	t	t.	:		y	.	y	V	(x	x	x		%)			

where:

- ^ indicates the selected actuator; it is displayed on the *Out* line instead of the actuator number;
- f indicates the forced/fixed state of the actuator; it is activated pressing the  button; use the keys   to shift on the selected actuator; it can be assume the F value if the forcing is active or it isn't displayed if the actuator follows the programmed actuation logics.

Use  key to shift to diagnostic window type 5 (this window appears only if the configuration is provided to GPRS):





0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
G	S	M	:	M		G	P	R	S	:	P								
C	S	Q	:	c	c		E	r	r	:	e	e							
C	O	n	N	.	:	n	/	t											

where:

- m: Status of connection to GSM network (0 “not connected”, 1 “connected”);
- p: Status of connection through GPRS protocol (0 “not connected”, 1 “connected”);
- cc: quality of GSM signal (0 “no signal”, 31 “top signal”); the value lower than 15 doesn't allow the GPRS use;
- ee: code of GPRS connection failure (0 “no error”);
- n: number of right connections (from last statistics reset);
- t: number of total connections (from last statistics reset).


In order to safeguard the modem memory the statistic is saved every 20 connection attempts and only if modem isn't powered through actuator.

For further details about the understanding of these window's data see §5.6.4

During the display of this window press key  to arrange immediately, for diagnostic purposes, the connection with the operating centre. Otherwise press  to start the modem reset immediately  and  to cancel its connections statistic.

The window shows the status of the GPRS modem only if the instrument's configuration considers its use (only for serial port 1); at the opposite the display goes to the next diagnostic window.

In case the configuration considers the use of the GPRS modem, but the modem is switched off, the window shows only the note : “Modem GPRS: off”, alternatively to the above mentioned information.

The key  shows the extra information about the GPRS connection, that are useful for LSI LASTEM technical staff:


- S: number of seconds remaining to the next status change of the procedure for the GPSR connection management;
- R: remaining tries against error, before current operation log out ;
- FSM: present status of the status machine.

If R-Log has been programmed for modem switch on and switch off, when the modem is switched off (only in this case), the following message is displayed on the first line of the window:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
M	o	d	e	m		G	P	R	S	:		O	f	f					

If R-Log has been programmed to operate with GPRS modem, but there are some communication problems between two devices or with the GPRS network, the diagnostic window type 5 could show the following message:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
M	o	d	e	m		G	P	R	S	:		N	.	C	.				


Use  key to shift to diagnostic window type 6 (this window appears only if actuation algorithm are configurated):

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
A	l	a	r	m	:	m	m		m	m	.	.	.						
																		m	m

where:

mm: number of measure in alarm conditions, according to the status obtained by the realization algorithm associated with the same measure. The window can display max 26 alarm measures (the first in the list order).



Use  key to shift to diagnostic window type 7:

0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
Z	i	g	B	e	e				S	N	:	n	n	n	n	n	n	n	n
S	t	a	t	:		a	a	a	a	a	a		(y	x)			
P	A	N	:	p	p	p	p	p	O	C	:	o	o						
S	S	:		s	s	%		R	C	:	r	r							

where:

- ZigBee indicates the radio type mounted on the instrument (if “ZigBee#”, the radio is a low power version);
- *nnnnnnnn* indicates the serial number of ZigBee radio, corresponding to the lower part (less significant) of the full address of ZigBee network;
- *aaaaaaaa* indicates the radio connection status and it can take these values:
 - *Undef.*: indefinite status;
 - *Init...*: radio module in progress of initialization;
 - *Init OK*: radio module initialized successfully;
 - *Conn OK*: status of correct connection to the PAN;
 - *No Conn*: status of not established connection to the PAN;
 - *Fail*: radio module absent or not working.
- *y* indicates the radio power on status in a Slave datalogger and it can take these values:
 - *W*: wake status, radio is active;
 - *S*: sleep status, radio in low power standby mode.
- *x* indicates the modem status; it can take these values:
 - *0*: the modem has undergone a hardware reset;
 - *1*: the modem has undergone a reset caused by a watchdog (blocked program);
 - *2*: connection to the PAN occurred;
 - *3*: status not related to the PAN;
 - *6*: master modality activated.
- *ppppp* indicates the programmed identity number of the PAN (a value from 1 to 65000);
- *oo* indicates the channel number used by the radio network (typically from 1 to 14; up to 16 if using low power radio); it appears only when PAN is connected;
- *ss* indicates the signal strength and it is expressed as a percentage; it is the received power of the radio signal of the last instrument directly connected (not include the intermediate repeaters);
- *rr* indicates the numbers of slave type nodes remaining available for connection to the instrument (maximum 10 for coordinator-master and maximum 12 for repeater-router); it is effective only for coordinator-master and repeater-router.





Press the  key in order to show the second diagnostic window dedicated to ZigBee:


0	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19
Z	i	g	B	e	e		F	S	M	:	f	f							
R	S	T	:	r	r		U	C	R	:	u	u							
C	E	:	c	c		I	C	:	i	i		I	P	:	p	p			
R	T	:	t	t		D	F	:	d	d		(x	/	y)			

where:

- *ff* indicates the number of current status of the main program for managing the radio driver;
- *rr* indicates the number of occurred reset of the radio caused by the driver;
- *uu* indicates the number of answer of the radio that doesn't match with AT command sent;

- *cc* indicates the number of management errors of the radio module of the received command;
- *ii* indicates the number of invalid commands;
- *pp* indicates the number of invalid command parameters;
- *tt* indicates the timeouts waiting the answer from the radio module to the management driver;
- *dd* indicates the number of failed transmission attempts;
- *x* indicates the status of the last radio data transfer; it can take these values:
 - 0: correct transmission;
 - 2: failed transmission;
 - 22: address node not valid;
 - 33: failed confirmation of network acknowledge;
 - 34: no network connection;
 - 35: addressing of network node obtained independently;
 - 36: address of network node not found;
 - 37: network track not founded;
 - 116: data size too high for transmission.
- *y* indicates the status of detection of network address of the remote node; it can take these values:
 - 0: no *discovery* operations was required for data transmission;
 - 1: a *discovery* operation was required to data transmission;
 - 2: a *route discovery* operation was required to data transmission;
 - 3: an *address* and a *route discovery* operation was required to data transmission.

Remember that by pressing at the same time of  and  buttons in ZigBee diagnostic windows resets the radio modem (press twice consecutively in a short time) or only statistics (press once).

Key  to return again to the product presentation mask.

4.3.5. Shutting off the display

Shut off the display allows to save about 25 mW. It is therefore important to keep the display shut off whenever possible when the instrument is running on batteries with or without sun panels.

To shut off the display you can use the keyboard (see §4.2) or use program *3DOM* to have it shut off; to do so, modify parameter "Characteristics – auto display shut off". By setting this parameter to *Yes*, R-Log will shut off the display after the keyboard remains inactive for three minutes.

4.4. Reset button

On the bottom panel, together with inputs for sensors (with terminal board or mini-din connectors), there's an hole for access to *Reset* key; the latter is only accessible through a pointed object (such as a straightened paper clip) and allows to instrument the reset of its possible block condition. Inside

memory will remain the instrument configuration, processed data and date and hour of system; on the contrary all statistics will be reset.

4.5. LED's

On frontal panel of R-Log there are four bright indicators (LED) that display the operation state of instrument (OK/ERR), the communication from and to external devices (Rx/Tx) and the charge-condition of inner battery.



See below a detailed description of each state.

Status indicator (Err, Ok)	Communication indicator (Rx/Tx)	Battery (Batt)	Description
Off	-	-	Instrument switched off, or low feeding voltage (min. 3,5 Vcc)
Single quick (5 s) green blinking	-	-	Instrument switched on. Regular operation (acquisition mode)
Three quick green blinkings	-	-	Instrument switched on. The instrument operates in reconfiguration mode
Five quick green blinkings	-	-	The instrument is on. The instrument operates in quick acquisition mode
Red slow blinking	-	-	Instrument switched on. An error occurred. The number of blinkings indicates the type of error (see §6.3)
-	Off	-	No current communication
-	Blinking	-	Instrument is receiving data from the serial communication lines or from radio
-	-	Switched off	Instrument feed from external; battery is fully charged
-	-	Switched on	Battery is charging

Interval between blinking cycles: 5 s.

Interval between blinking cycles in Slave datalogger with low consumption mode: 15 s

Quick blinking period: 125 ms

Slow blinking period: 375 ms.

5. R-Log additional features



R-Log is an instrument designed to acquire, elaborate and store measures from sensors connected to its analogue, digital and serial inputs. On the base of the acquired data it can operate field actuations according to the programmable algorithms.

5.1. Starting and running the survey

Note: The first thing to do before proceeding with the survey is to check the date/time displayed by the datalogger; in case of discrepancy with the current date/time, update them (see §4.3.1.7).

R-Log performs the measurements and processings as they always belong to only one survey. The execution of survey is made in automatic way, few seconds after the starting of instrument.

R-Log is a mono-survey instrument, but offers to the user the possibility to mark the time of surveys; in other words it allows to fix manually, for one survey, the starting and ending time of data acquisition. These information are shown by LSI LASTEM application programs in order to make easy the selection of data included into given period of sampling.

The starting time, to consider data, is fixed pressing  and  keys at the same time; an univocal numeric value n will identify the selected slot; n will be increased every time a new slot for acquired data consideration is fixed. The reset of survey number is made only modifying configuration from PC or starting new auto-recognition procedure.

This feature is complementary to datalogger standard acquisition mode; R-Log will always continue to acquire and store all data in the only survey.

After changing the configuration, a new survey will be started and all stored data in memory will be deleted. The PC's application software is used for a complete data management procedure.

5.1.1. Automatic recognition of probes

R-Log dataloggers complete with inputs with mini-din connectors (series ELR510) are equipped with function for auto-recognition of probes made by LSI LASTEM.

The inner library of R-Log allows the instrument its auto-configuration according to LSI LASTEM sensors physically connected to datalogger (for its operation see §4.3.1.3); the library assigns the respective measurements (acquired and calculated) to each specific recognized sensor and in this way it automatically setup the configuration with more suitable parameters.

The measurements configured in this library (programmed thanks to experience of LSI LASTEM engineers) have acquisition times and other management modes in accordance with type of measured quantity. It's however possible to modify the configuration parameters (processing time, acquisition rate, name of the measurement ...); in order to carry-out this modification, follow the instruction below:

- 1- open 3DOM software;
- 2- download configuration from instrument (see §4.4.8 of SWUM_00339 manual);
- 3- modify the interesting parameters (see §4.4 of SWUM_00339 manual);

- 4- save the configuration and re-send it to instrument (see §4.4.7 of SWUM_00339 manual). The programming of new configuration produces the erasing of all data in the instrument memory; for this reason it's recommended to transfer data on PC, through same program, before you setup new configuration.

In order to allow the instrument the right acquisition, the acquisition rates of measurements must be lower or equal to elaborated rate. During auto-configuration function, one or more measurements, having acquisition rate higher than rate setup for processing, can be generated in R-Log; to avoid this possible error, R-Log has been equipped with function of automatic correction that restores the acquisition rates (if they are higher) equal to elaborated rates automatically.

During auto-recognition function could be pointed out some errors; here below the codes and their relevant meanings:

1	GR_CISS_Error	In the instrument the library of sensors includes the description of sensor or non-valid measurement
2	GR_EmptyLibrary	The instrument doesn't include a valid library of sensors
3	GR_ParamNotFound	The measurement doesn't include the placeholder for demanded parameter
4	GR_UndefinedProbe	The connected sensor hasn't been defined in library of instrument
5	GR_UndefinedMeasure	The connected sensor generates a measurement that hasn't been defined in library of instrument
6	GR_DepMeasureNotFound	One calculated measurement requires one dependent measurement (acquired or calculated by same or other sensors) that hasn't been generated
7	GR_DepMeasureMarkerNotFound	The tag for dependent measurements, included into measurement definition, hasn't been found
8	GR_MeasureUpdateRateNotFound	The tag for the update rate, included into measurement definition, hasn't been found
9	GR_SensorCalibrMarkerNotFound	The tag for calibration value, included into measurement definition, hasn't been found
10	GR_UndefinedProbeCalibrParam	One connected sensor requires the definition of calibration value that hasn't been defined in standard quantities (*)
11	GR_UpdateActiveConfigError	Memory writing error during update of instrument current configuration

(*)Algorithm for search of parameters for sensor calibration: if there's only one defined parameter it's always used for every input (also more than one); if there're several defined parameters, they must correspond with specified input; in other cases, error. The calibration parameters can be defined at the end of list of standard quantities.

The function of probes auto-recognition is a local setup, in other words the configuration is set-up automatically and just in the instrument where this function is required. In case there's a network of R-Log dataloggers, it's necessary the use of 3DOM software in order to update the configuration of master instrument and acquire properly all quantities coming from other acquirers by radio.

5.1.2. Checking the power supply voltage

During the survey the instrument checks the power supply voltage constantly, in order to grant its enough to the inner functions. The check isn't arranged to grant the operation of the outside devices (modem, sensors that need power supply, etc.) that already couldn't work regularly.

The check happens every minute: the instrument acquires and checks the power supply voltage; it must be 3,5 Vdc at least. In case it isn't enough, R-Log closes the survey in progress and stands by. During the stand-by it checks every minute the power supply voltage, until it'll reach the minimum value of 3,7 Vdc. The note "Power low" (instead of the measures list) shows the stand by condition looking forward the right power supply. When the right voltage comes back, the survey's re-opened and the standard operating modes re-start.

Note: the power supply voltage is acquired and checked, even if the instrument's configuration doesn't include (among the programmed measures) the corresponding acquisition measure of the battery's voltage. In order to store the feeding values (surveyed during the survey), the battery voltage measure must be programmed (like proposed by *3DOM* warning message).

5.2. Acquisition and calculation of measures

It is possible to acquire measures of sensors with a rate from 1 second to 12 hours; this allows a better representation of both fast-changing quantities (wind speed), as well as slow-changing ones (air temperature). It is possible to acquire as many as 10 measures per second.

For sensors which have to be powered, R-Log is fitted with own outputs with switched power supply (*actuators*); see §3.3.3.

The instrument is able to calculate quantities deriving from measures that sample the signs from the inputs: R-Log is fitted with a calculation library dedicated to environmental applications, also able to carry out useful mathematical functions (see §6.2). R-Log is able to acquire and calculate up to a maximum of 50 total measures. Furthermore it's possible to program some calculated measures according to the data generated by other calculated measures.

Program *3DOM* can be used to program the measure sampling sequence: the program interface allows choosing the sequence in which the measures are displayed; hence which is the sampling sequence taking place during the survey. With *3DOM* you can also set automatically the measure sequence according to their acquisition rate (from the shortest to the slowest). This is important to gather measures which, during the acquisition process, must be sampled within the shortest intervals possible among each other (if they are allocated to only one actuator).

If configured with the same acquisition rate, the analogue channels following the first one will be sampled about 80 ms after the previous channel. This means that as many as eight analogue channels are sampled within a total time of 400 ms. Digital channels are instead sampled in about a flash. In case of application with a sensor with two measures that requires an actuator, remember that the probe must be connected on 1-2 inputs (actuator n.1) or on 3-4 inputs (actuator n.2).

5.2.1. Acquisition from sensors with serial or radio output

R-Log is able to acquire measures from sensors connected through radio or serial port 1. R-Log actually supports LSI LASTEM CISS protocol (owner communication protocol designed by LSI LASTEM and available in all sensors with microprocessor; medium for via radio reception by means of LSI LASTEM DEC301 outside receiver or ZigBee inside radio).

The use of CISS protocol allows the connection to one or more LSI LASTEM sensors (even if they have different model); every configured sensor must have an univocal network address as regards other sensors connected on the same communication line. The connected sensors' quantity determines the sampling minimum slot of the instrument: consider usually 3 sensors per second (the minimum acquisition time with 20 sensors must be set up 7 seconds at least). It is important that the acquisition time programmed in the instrument through *3DOM* application (shown by the parameter "*Update rate*") must be the same of the spontaneous communication rate programmed by the sensor through *LSM* program.

The sensor can be programmed to repeat the message communication several times, in order to improve R-Log reception reliability; one repeat is usually enough (so the message has been transferred twice); before the arrangement of several repeats, take into consideration: the power autonomy of the sensor (if it is powered on through battery), and the further generated traffic. This can make worse the reception reliability (depending on the connected sensors and the programmed acquisition slots).

The acquisition of messages from the sensors happens as follow: *3DOM* application programs one measure for each quantity that has been acquired and transferred inside the message by the sensor; every measure is correlate to the respective sensor's quantity by means of the sensor's network address, that specify: the origin sensor (in *3DOM* it's the parameter *Probe ID*), and the ordinal number of the quantity inside the message; for example in case of LSI LASTEM mod. DME811 sensor (programmed by network ID = 5) can be programmed up to 5 different measures, which correspond to the quantity: *air temperature*, *relative humidity*, *surface temperature*, *temperature*, *temperature*. The quantities number and their programming order aren't binding (for example it's possible program *surface temperature*, *relative humidity*), but the quantities sequence must have the same order like in the message; in the previous example two R-Log measures must be programmed as follow:

<i>Quantity</i>	<i>Network address</i>	<i>Measure index</i>
Surface temperature	5	3
Relative humidity	5	2

3DOM application automates some configuration operations by means of its sensors library: for example to program the LSI LASTEM mod. DME810 sensor press key *Add* from the measures list window and then select the relevant code from the available sensors list: the application knows that sensor is serial type and so it requires the sensor's network ID; in succession the procedure generates all needed measures for the sampling of the sensor in the right way.

5.2.2. Acquisition from thermocouples

R-Log is able to acquire the signals from many types of thermocouples. R-Log uses the internal temperature value as reference of value of the cold junction.

In these cases it is necessary to program, by means of the *3DOM* program, the measure of the internal temperature. The measure of the internal temperature must precede, in the measure sequence, all measures of quantities which use such reference.

5.2.3. Details about the measure acquisition process

The sampling of the signals produced by the sensors connected to the instrument's terminal board takes place according to the following logical procedure:

- 1) Measurement of the electrical signal based on its type (voltage, resistance, frequency, etc.) and its digital conversion into a 16-bit numeric value; the physical type of the sensor is programmed through parameter *Electrical measure type*;
- 2) Data validation: during this operation the value is limited within the scale values allowed by the physical type of measure;
- 3) Possible thermocouple value correction through the measure of the cold junction temperature (internal temperature of the instrument);
- 4) Linearization of non linear signals based on the setting of parameter *Linearization type*; the linearization may also take place through the setting of a polynomial function whose factors are specified until 10° degree (section *Linear parameter* of *3DOM*);
- 5) Recalculation of the value according to numeric parameters defined in section *Parameters*:
 - Computation of the measured quantity through the defined initial and final scale values;
 - Application of the calibration factor of the specific used sensor (radiometers, rain gauges, etc.);
 - Selection of the logical state with reference to the analogue signal thresholds;
 - Measure validation after processing accompanied by error indication if greater (by 0.5%) than the limits set in output; wind direction and relative humidity are excluded;
 - Linearized quantity control: the instrument produces null output when receiving null input;
 - User scale limits.

All above parameters are indicated in section *Measure properties* in the measure modification window of program *3DOM*.

5.2.4. Acquisition from status signals

R-Log is able to acquire different types of digital status; they have to be configured for their connection to input number 5.

There are 3 different signal types: frequency signals, digital status and counters.

The default configuration of datalogger has been made for:



give logical status = 1	In case of short circuit or 0 V
give logical status = 0	In case of opened contact or 4 V

The sensors with status output that produce voltage (i.e. they aren't pure contacts "open/closed") but with variable voltage according to measured status, can be connected to R-Log through one diode; in this way every connection is always the right one, apart from the output voltage (no divider is required). The anode of diode must be placed on clamp F of terminal board's entrance and the cathode towards the sensor.

For long acquisition it's advisable power on externally the datalogger.

5.2.5. Fast acquisition mode of the measures



For diagnostic purposes, R-Log is equipped with a function that allows the acquisition of all sensors connected to its inputs at maximum speed (excluding the sensors connected to the serial port or received through the radio).

In order to activate the fast acquisition mode press key  and  at the same time. In this mode, the acquisition rate becomes 1 second every 10 measures and LED of the instrument will show a particular flashing: 5 green quickly blinkings.

Please note that in this condition the instrument:

- Acquires all sensors and recalculates all measures with minimum rate of 1 second every 10 measures;
- Keeps switching powers used to power the sensors permanently on;
- Uses up much more energy;
- Produces elaborations at the programmed rate; uses a higher number of samples compared to the normal conditions;
- In this mode the green *Ok* LED blinks 5 times every 5 seconds.

In order to set the instrument back to its normal mode, switch it off and on again without however

pressing key  and  at the same time; as indicated above.

5.2.6. Activation of sensor control

Function valid only for R-Log with connectors (ELR510 model)

The instrument acquires the sensors rapidly thus using up little energy; hence increasing its operative autonomy.

When enabled, this function will replace the normal acquisition process. Avoid setting a check that is too frequent, especially in case of several fast rate measures. Do not use this function if you aim at saving as much energy as possible. For example, in case you have programmed 8 measures with 10 seconds acquisition rate, set the Probe check rate to 1 minute or higher.

5.2.7. Details about calculated measures

If the instrument has been programmed to process one or more calculated measures, the logical process will be the following:

- 1) Acquisition of all primary measures that allow the estimate of calculated measures; a calculated measure can be a primary measure for new calculated measure;
- 2) Collection of the value of primary measures; if at least one of these values is found to be in error, its calculated measure will also be indicated in error;
- 3) Collection of the value of standard parameters, whenever used in the calculation; the value of these parameters is decided during the configuration process and cannot therefore be modified during the survey;
- 4) Execution of the estimate;
- 5) Allocation of the calculated value to the measure's instant datum.

The acquisition rate of a specific calculated measure is set by *3DOM* so that it may correspond to the acquisition rate smaller than the calculated measures it depends on.

5.3. Measure elaboration

For each acquired or calculated measure it is possible to obtain statistical elaborations at a time base from 1 second to 12 hours. The selected elaboration base is common to all quantities.

Just as with the acquisition process (see §5.2), the elaboration process, too, evaluates the time of the built-in clock as a multiple of the elaboration rate in order to determine the moment in which the elaboration of the statistical data begins. For example, if the elaboration rate were to be set to 1 hour and 30 minutes, and the current time were 15:24:01, the following elaborations would take place at 16:30:00, 18:00:00, 19:30:00, etc.; the elaboration uses all instant data acquired or calculated in the chosen elaboration span.

The available statistical elaborations are:

- Arithmetical calculations
 - Instantaneous value
 - Mean
 - Minimum
 - Maximum
 - Standard deviation
 - Total
 - % valid data
- Vectorial calculations specific for anemometric quantities
 - Prevailing direction
 - Resulting direction
 - Resulting speed
 - Direction standard deviation (sigma-teta)
 - Calm wind percentage

It is not possible to combine arithmetical and vectorial calculations for each specific measure.

Program *3DOM* can be used to program the elaboration parameters.

The instrument doesn't have infinite process capacity: it depends on the number of acquired and calculated measures, on programmed actuation algorithms, on processings configured for each measure, and on continuous communications between the instruments and the outside devices; all these parameters cannot be programmed at the same time up to their available maximum, because the instrument could have wrong operation. The instrument has right operation in the following heavy duty condition:

- All analogue and digital inputs configured with measures of resistance, tension, state and frequency (4 kHz); every input sampled with 1 second rate;
- Six measures configured like measures calculated with several algorithms;
- The remaining measures (up to 50 measures) sampled by LSI CISS sensors with transmission every 10 seconds;
- Each measure is elaborated every 30 seconds with minimum, medium and maximum value statistics and standard deviation;
- All 10 active actuation logics configured with different algorithms and by the use of acquired and calculated measures;
- Serial communication constantly activated on both communication ports at maximum bit rate.

It's possible to *relax* some parameters in this configuration (for example the measures total number or the active actuation logics number) in order to obtain best performances in other cases (for example the measures activation slot)

5.3.1. Vectorial calculation specific for anemometric quantities

Average direction (prevailing)

$$Dm = gra(atan2(\Sigma Sin(rad(Dir)), \Sigma Cos(rad(Dir))))$$

Resulting average direction

$$Dmr = gra(atan2(\Sigma (Sin(rad(Dir)) \cdot Vel), \Sigma (Cos(rad(Dir)) \cdot Vel)))$$

Resulting average speed

$$Vmr = \frac{\sqrt{(\Sigma Sin(rad(Dir)) \cdot Vel)^2 + (\Sigma Cos(rad(Dir)) \cdot Vel)^2}}{n}$$

Direction's standard deviation (sigma theta)

$$DevStandDir = gra \left(asin \left(\sqrt{1 - \frac{(\Sigma Sin(rad(Dir)))^2 + (\Sigma Cos(rad(Dir)))^2}{n^2}} \right) \right)$$

Calm Percentage

$$\text{Calm} = \frac{\sum_{i=1}^n \text{Calm}_i}{n} * 100$$

Where:

Dir = instantaneous value of wind direction (0 - 360 °)

Vel = instantaneous value of wind velocity (m/s)

gra = conversion of an angle from radians to degrees

rad = conversion of an angle from degrees to radians

Calm = 0 in case of not calm wind velocity (< 0.3 m/s), otherwise 1

n = number of considered valid original data (no error)

5.4. Storing elaborated data

R-Log stores the calculated statistical processings (elaborated data) in the 2-MB internal memory; a part of this memory (128 kB) is used for configuration information and other internal information; the real capacity is therefore slightly less than the total storage capacity.

Data storage takes place circularly; once the memory is full, the new data will replace old ones.

By sending a new configuration to the instrument, all data so far stored will be cancelled; this happens because the new configuration information might be potentially not in tune with them, and the PC might misinterpret them.

The instrument stores the data in the internal memory only when a *data page* is full; the dimension of this page equals to 256 bytes; therefore the instrument might store the data only after several sequences of elaboration; this depends on the programmed elaboration rate, on the number of active measures and, for each one of them, on the selected elaboration elements. Please note that by switching off the instrument, some elaborated data inside the *data* page and those not definitely stored might be lost; for this reason we advise transferring the elaborations which have not yet been transmitted to the PC before switching off R-Log.

5.4.1. Memory autonomy

Depending on the chosen configuration (measures, types of elaboration for each measure and acquisition rate) the instrument will work somewhat autonomously as far as the maximum storage time is concerned, without replacing old data with new ones. The calculation of the time dimension of the maximum number of storable data is as follows:

$$A = K / (86400 / RE * NE)$$

where:

A = number of days of autonomy of the data memory;

K = value subordinate on the size of the used memory; for the 2 MB internal memory, *K*=386048;

RE = elaboration rate expressed in seconds;

NE = total number of elaboration's elements programmed for all measures.

5.5. Actuation logics

R-Log has got an actuation logics library. It's useful to switch-on all type devices (alarms, solenoid valves, motors) according the surveyed parameters in surrounding environment. The actuation logics are based on the instantaneous value of the acquired and calculated measures. They can be programmed up to 10 calculation algorithms, that use same or different logics. One or more algorithms can be combined in two differed modes , in order to switch-on the selected actuator:

- 1) All algorithms must be in alarm at the same time (AND logic);
- 2) One algorithm can be in alarm at least (OR logic).

The actuator's switch-on logic can operate according to *low power consumption* mode (the actuator is usually deactivated, and starts in case of alarm), or according to *safety* mode (the actuator is usually activated, and it shuts-down in case of alarm). Summarizing explanation in the table below.

<i>Operating logic</i>	<i>State</i>	<i>Actuator</i>
Low consumption mode	No alarm	Actuation output switched-off
	Alarm	Actuation output switched-on
Safety mode	No alarm	Actuation output switched-on
	Alarm	Actuation output switched-off

In case of one or several measures' error (for example owing to sensor breaking, no-scale acquisition, or disconnected cable) it doesn't modify the present state of the actuator piloted by the logic that uses the same measures.

The manual activation and deactivation of the actuators, performed by the user directly on the instrument from *Diagnostic window type 4* (see §4.3.4), does not affect the actuation logics. The actuation logic deactivates the actuator only if it has previously activated from the logic itself.

The actuation logic begins operation after the first activation of the actuator, even if this occurs temporally after the deactivation.

The programming of the actuation logics happen by means of the *3DOM* program according to two different phases:

- 1) Selection of the logics and their calculation parameters (section *Logics*);
- 2) Selection of the actuation outputs and their AND or OR modes correlation with the present logics (section *Actuators*); note that the same logic can be combined with other different logics several times, in order to switch over different actuators.

For R-Log with connector inputs (ELR510 series) may only be used the following actuation logics:

- Cyclic timer or with data/hour start/stop,
- Threshold value compare.

For R-Log with terminal board inputs (ELR515 and ELR516 series) may be used all the actuation logics available from 3DOM software.

5.5.1. Eolic alarm

(Logic used only for R-Log with terminal board inputs – ELR515 and ELR516 series)

The logic uses a wind direction measure to establish the condition of a wind position in a specified sector for a predefined time. Can be set:

- The measure that samples wind direction (degrees);
- The starting angle (extreme included) of the direction sector;
- The ending angle (extreme included) of the direction sector;
- The continuous permanence time of the wind direction inside the defined sector in order to detect the alarm condition;
- The continuous permanence time of the wind direction outside the defined sector in order to detect the end of the alarm condition.

Both times can be set from 0 seconds to 12 hours; if both times are set to zero, the wind direction in or out conditions inside the sector are immediately detected.

It is possible to join this logic with an another logic of threshold bypass type (see §5.5.5) applied to a wind speed measure: in this way it's possible to refine further the alarm activation (i.e. to activate the alarm when the wind is more than 5 m/s for at least 3 minutes and inside the *Est* sector of 45 degree for at least 1 minute).

5.5.2. Evaporimeter filling

(Logic used only for R-Log with terminal board inputs – ELR515 and ELR516 series)

The logic uses a water level probe inside the evaporimeter to establish the need for filling up it. Can be set:

- The measure that samples the water level;
- The filling-up start time (it's recommended the automatic filling-up programmed in the morning, before the sunrise, in order to avoid temperature changing that could alter the evaporation measure);
- The maximum filling-up time, useful to avoid flooding in case the water level sensor is broken or surveys a wrong measure;
- The maximum water level that determines the stop of the filling;
- The minimum water level, under that the need for the evaporimeter filling-up (at the defined time), is detected. To obtain the right evaporation keep the evaporimetric basin always filled. Therefore set the minimum level like the maximum level, because, in case of too low water level, the shade of the walls on the surface of the water doesn't allow the right evaporation in the morning and in the evening.

5.5.3. Start precipitation alarm

(Logic used only for R-Log with terminal board inputs – ELR515 and ELR516 series)

The logic uses a measure connected to a rain gauge to detect the start precipitation conditions. Can be set:

- The measure that samples the precipitation;

- The minimum time T1 after the first precipitation detection by means of the rain gauge (instantaneous value > 0);
- The minimum time T2 that must pass after the precipitation detection, meanwhile no precipitation is detected (no precipitation detection by the inner sensor of the rain gauge), to determinate the end precipitation condition;
- The minimum precipitation quantity that determines the start of the precipitation condition.

The alarm condition is detected when T1 time is passed after the first precipitation detection (and it's raining again), or is reached the specified rain quantity; anyway if T2 time is passed without any precipitation detection, the system goes to no alarm condition.

5.5.4. Flood alarm

(Logic used only for R-Log with terminal board inputs – ELR515 and ELR516 series)

The logic uses a measure connected to a rain gauge to detect flooding conditions. Can be set:


- The measure that surveys the precipitation;
- The maximum precipitation quantity in a defined period;
- The minimum precipitation quantity in the same period;
- The length of alarm or not alarm period.

The alarm condition is detected when, within the specified period that starts from the first precipitation event, the maximum precipitation quantity is exceeded; from alarm condition beginning or at the end of the first period, are managed new periods and for each of them the rain totalization starts from zero; for each new period, if the rain fall quantity returns below the specified minimum value, the system returns in no alarm condition.


5.5.5. Threshold value compare

The logic detects value overflow or underflow by one or more measures (both univocally and simultaneously). To the threshold values can be applied a further hysteresis value; this can avoid continuous alarm state changes in case the measure value moves nearly around the threshold value. The comparison logics are the following:


- *Greater than*: alarm in case the measure value is greater than the threshold value added to the hysteresis; return to no alarm condition when the measure value is lesser than the threshold value minus the hysteresis; i.e. threshold value=4.0 and hysteresis=0.2 (alarm in shaded area):




- *Lesser than:* alarm in case the measure value is lesser than the threshold value minus the hysteresis; return to no alarm condition when the measure value is greater than the threshold value added to the hysteresis; i.e. threshold value=4.0 and hysteresis=0.2 (alarm in shaded areas):



- Included:** alarm in case the measure value is greater than the minimum threshold value and at the same time lesser than the maximum threshold value; return to no alarm condition when the measure value is lesser than the minimum threshold value or greater than the maximum threshold; hysteresis is used like shown in the following example: threshold values=2.0 and 4.0 and hysteresis=0.2 (alarm in shaded areas):



- Excluded:** alarm in case the measure value is lesser than the minimum threshold or greater than the maximum threshold; return to no alarm condition when the measure value is greater than the minimum threshold and at the same time lesser than the maximum threshold; hysteresis is used like shown in the following example: threshold values=2.0 and 4.0 and hysteresis=0.2 (alarm in shaded areas):



Threshold value compare algorithm can be applied to one or more measures; in the last case all used measures must be consecutively ordered (the list cannot include any measures that shouldn't be considered in the comparing), because the logic programming needs the ordinal numbers of the first and last measure. This logic can be programmed to detect the alarm condition only if all measures (that belongs to the selected set) or alternatively only one of them are contemporary over the stated limit.

Furthermore it's possible specify a minimum time (during that the measure stays over the set threshold value) to enter into the alarm condition and a minimum time to exit from the alarm condition. It can be done applying the algorithm to only one measure (i.e. it isn't possible to program the minimum times if two or more measures have been selected).

5.5.6. Timer

The timer logic allows: to activate or deactivate the actuator in two different times in the day, or alternatively to define a timed cycle on/off status. Can be set:

- Timer type (cyclic or time);
- The power-on delay after the initial instant determined by the cycle;
- If timer type is cyclic, the on and off statuses duration; the first cycle starts in the day time when the instrument clock time, divided by the sum of the two durations, returns zero as remainder; in this way the first cycle starts in a precise day time, not at any moment (i.e. if has been programmed 15 minutes *On* period duration, and 45 minutes *Off* period duration, the first cycle starts at the first hour and zero minutes after the instrument survey is started); the next cycles happen at set on/off times.
- If timer type is not cyclic, the day time of power on the actuator, and the day time of power off.

This logic is joinable with other logics in *AND* mode, for example to allow the alarms activation only in specified hours of the day.

5.5.7. Snow level alarm

(Logic used only for R-Log with terminal board inputs – ELR515 and ELR516 series)

The logic detects excessive snow fall conditions in indeterminate time period; the alarms condition stays active during programmable time; at the end of this alarm period the snow level starts to be counted from the actual level; if during the alarm condition the snow level decreases (owing to melting or autocompression), the snow level start value (used like reference value in the next delta calculation) updates accordingly. Can be set:

- The measure that samples the snow level;
- The maximum delta (centimeters), over that the system goes in alarm condition;
- The alarm duration before its automatic reset.

5.5.8. System error

The logic sets an alarm when the instrument detects and internal malfunction. All errors are detected as specified in §6.3.

5.6. Communication modes

R-Log has got different communication devices, that can be used to extend the connection capacities and modes to systems for data collection:

- RS232 serial port: one port (equipped with a double connector for applications where you want to interface R-Log with a DCE device or a DTE device) for each R-Log datalogger, it's available for all models excluding ELR516; (see §5.6.1)
- ZigBee radio: this technology is available for each R-Log in order to allow the reception of LSI LASTEM cordless sensors and configure nets of datalogger that communicate together; its operating frequency is 2.4 GHz, it's inside the instrument; (see §5.6.2)
- RS232/RS485 serial converter DEA504: it lengthens the cable for serial connection up to 1 km or more; the connection speed is function of covered distance; you have always to install a couple of line drivers: one of them must be connected to RS232 serial port of R-Log, the other one must be connected to PC; connection scheme: DISACC5584a.
- RS232/Ethernet converter DEA550: it can be used with instruments equipped with RS232 serial line; it allows the communication using Ethernet LAN/WAN net, so virtually of any distance; particularly ELR516M model has this converter built-in; (see §5.6.3)
- GSM DEA714-DEA715 Modem: it connects to remote devices by means of GSM network; connection rate 9600 bps; it can be connected to serial port of R-Log; DEA715 modem (using actuated signals from datalogger) can be used for application with sending of SMS in case of alarm state; connection schemes: DISACC4852b (DEA714), DISACC4978a (DEA715).
- GSM/GPRS DEA717-DEA718 Modem: it uses the GPRS data packet transmission to allow the continuous transmission (with transmission rate according to choice) and rate according to traffic; it can be connected only to serial port 1 of R-Log; connection schemes: DISACC5416 (DEA717), DISACC5416a (DEA718).
- Radiomodem devices DEC010/5/8/9: they permits communications at long distance (some kilometers) operating on VHF 169 MHz and UHF 868 MHz frequency ranges; they don't require government concession; user's manual: INSTUM_00757.

5.6.1. RS232 serial port

R-Log communicates with PC through 9 poles male/female not-reversing serial cable.

R-Log implements a particular system for activation of communication port: after 8 seconds of no-reception or no-transmission, the consumption goes down to first saving; after further 22 seconds the consumption goes down to minimum values. If you don't query R-Log continually, it communicates using very low energy consumption.

Every instrument uses a net address, that consists of a number with value included between 1 (default) and 200. If the instrument has been connected through network (radio) with other instruments, it's necessary to modify the address. The PC programs use the address of each definite instrument to select R-Log for sending of communication messages.

R-Log is compatible with LSI LASTEM sensors family equipped with CISS protocol. So it can be used in existing networks of LSI LASTEM cordless sensors, that make reference to R-Log acquisition instruments, Babuc ABC or receivers connected to PC directly. In this case R-Log can

be programmed for spontaneous transmission, through serial port, of instantaneous data of acquired and calculated measurements; this parameter is available in mask for modification of 3DOM communication parameters, and it's called *automatic transmission rate*. In this mode R-Log results like LSI CISS multi-channel sensor and so its measurements can be received by an other R-Log.

A GSM or GPRS telephonic modem can be connected to serial port; in this case, if you want reduce the energy consumption of modem, it's possible follow the instruction of §5.6.4. Anyway, if removed appropriately, the use of a modem doesn't block the local communications with R-Log, connecting PC to serial port of instrument directly.

Note: in case PC has to execute any communication with the instrument through serial port, and the latter has been programmed for spontaneous transmission of data at frequent rate (intervals shorter than 10 s), PC could meet some difficulties to start first communication; so it could be necessary to start operation several times. After the reception of message, R-Log deactivates spontaneous communication for 1 minute, in order to make further communications with PC easier.

R-Log manages the flow check according to two choices (none, only RTS) that can be selected through configuration from 3DOM (see §4.4.4.3 of manual SWUM_00286 inside DVD of LSI LASTEM – MW6501 products).

<i>Protocol / Device</i>	<i>Function</i>	<i>Com1</i>	<i>Com2</i>
(Owner) LSI LASTEM CISS	Transmission of instantaneous values of the measures (in polling mode or spontaneous transmission)	X	X
	Transmission and reset of elaborated data from memory	X	X
	Transmission of instrument registry information	X	X
	Transmission and setting of configuration parameters (measures, elaborations, communication, actuation logics, etc.)	X	X
	Transmission and reset of diagnostic information updated in real time or stored in memory (system log)	X	X
	Transmission of actuation logics state (allarms) or digital outputs state (actuators) (*)	X	X
	Transmission and setting of system date/time	X	X
	Setting of digital outputs (actuators)	X	X
	Acquisition of instantaneous values sampled from sensors with LSI LASTEM CISS protocol	X	X
	Management of system commands (start/stop survey, memory format, instrument reset, etc.)	X	X
PSTN/GSM Modem	Available in transparent mode with any protocol	X	
GPRS Modem	Support to communication on TCP socket (encapsulation of the only owner LSI LASTEM CISS protocol), or with FTP protocol (transmission of elaborated data in binary format)	X**	
ZigBee	All commands provided in owner LSI LASTEM CISS protocol		X
RS232/Ethernet Converter	Available in transparent mode with any protocol	X	

(*) Function obtained by programming of appropriate calculated measures

(**) Function available only for ELR515 and ELR516 serie models

5.6.2. ZigBee Radio

All R-Log are equipped with radio with ZigBee protocol (international standards IEEE 802.15.4) and they are particularly suitable for applications where devices interact together through on average low temporal frequency and transporting not big quantities of data (packets consisting of some hundred bytes); they're excellent for network connection of a lot of devices (up to 200) that don't need non-stop communication channels, but need the possibility of data swap only on request. Besides, assigning different network addresses (PAN ID) during configuration, it's possible to obtain several nets for data acquisition that operate at the same time and parallel, without any interference problems.

Radio has been built in the instrument (on serial port 2) and the user can see only antenna coming out of top panel.

All R-Log are supplied with configuration set by manufacturer and already programmed according to one of three different possible modes (Master, Repeater, Slave) in order to obtain one network consisting of nodes that can communicate together.

- *Master* is the main device (always powered) for network management, and it's connected to host (PC) through direct connection or communication computer equipment; it's concentration point of all messages and surveyed data; in terminology ZigBee is called *Coordinator*.
- *Repeater* is continuously powered device that acts as measurement post and messages repeater inside network when they cannot reach master device directly; in terminology ZigBee is called *Router*.
- *Slave* is device that can remain in low consumption mode for a programmable time, minimizing consumptions; it's a standard device for applications with power supply through battery (also small capacities); in terminology ZigBee is called *End-Device*. (*)

Although it's usually used with wireless networks having limited coverage (i.e. that are able to cover areas with a range of about hundred meters), thanks to its peculiarity to configure the acquirers like repeater, it allows to monitor also a wide environment setting up a net of instruments that can communicate together without the use of special communication devices (such as transmitters and dedicated cordless repeaters). Besides, as there're several repeaters devices inside one network, the message can find valid alternative paths in order to reach the posting of communication in easier way or in case of failure of one network device.

The reasons of limitations connected to ZigBee application are the devices distance and physical obstacles (outer walls, walls and ceiling for indoor applications; buildings and trees for outdoor applications) that signal can find during way towards communication destination.

<u>Specifications</u>	<u>Performances</u>
Indoor standard range	up to 60 m
Outdoor standard range (as the crow flies)	up to 500 m
Diffused power	10 mW (+10 dBm)
Reception sensitivity	-102 dBm
Operating frequency band	ISM 2.4 GHz
Bit rate	250 Mb per seconds
Data throughput	up to 35000 bps
Number of channels	13
Transmission consumption (only radio)	170 mA
Reception consumption (only radio)	45 mA
Supported network topologies	Point-to-point, Point-to-multipoint Peer-to-peer, Mesh
Agency approvals	Europe (CE) ETSI

	United States (FCC Part 15.247) FCC ID:MCQ-XBEEPRO2 Industry Canada (IC) IC: 1846A-XBEEPRO2 Australia C-Tick Japan R201WW8215142
--	---


Here below the use conditions to follow for network implementation:

- every network must be equipped with only one Master device;
- use Slave instruments only for energy saving needs or in case they cannot be powered continuously;
- max number of cascade repeaters is 10, in other words the longer physical network distance must be covered using 10 sections;
- max 12 Slave devices can be directly associated with each repeater;
- max 10 Slave devices can be directly associated with each Master;
- total max number of devices for each network is 200 (the same for CISS protocol).

ZigBee network includes two main transmission modes:

- *spontaneous transmission* of messages from Repeater or Slave to Master using the network potentialities to reach Master position even if not directly connected with Master;
- *transmission on request* from Master ("polling"). Master connected with system for data call and storage (usually on PC) manages the communication examining (according to time settable from dedicated software) all devices in network; in case of subnets it'll be necessary to equip every subnet with a further dedicated receiver equipment in order to communicate with Master of network at the same time and without any interferences (in this mode the radios are always on, for this reason it's recommended to install radio repeaters, to give master several paths in order to reach the last devices; it's obvious that R-Log will not be able to operate in low consumption conditions).

5.6.2.1. Network set-up




If you want set-up new generic network, in other words in case you don't know if network will include Router and End Device devices both like distribution and number, we recommend to follow the points below:

1. Distribute all nodes known topologically (depending on density and reciprocal distance) covering the area of network use through computer devices that should be connected.
 - *Study the topology of place where you install the network devices.*
 - *Put the devices in place making attention to distance as the crow flies among respective nodes (remain within max range specified in specifications) and considering the obstacles on every distance (building works, furniture, vegetation...).*

2. Check that obtained distribution could have RF coverage also on longer distances (i.e. verifying that communications are carried out correctly in every intermediate distance)
 - *Check coverage in all distances using devices configured with quick acquisition/transmission rate. For each distance place the respective Router or End-Device device in installation point and go from that point away with Master device checking the coverage parameters and instruments connected to it (see §4.3.4 diagnostic mask 7).*
3. Evaluate the possible End Devices in case there are needs of powers through battery
4. Evaluate Routers devices (always powered) that allow the network arrangement with End Devices. These Routers can be other devices with measurement facility, or can be exclusively dedicated for collection and transmission of messages generated by End-Device devices and their members
 - *Check that all positions preset for Routers devices are equipped with 220 Vac or 12 Vdc power supply.*
5. Value the redundancy obtained with all present Routers considering the possibility to have multiple paths suitable for data transmission. In practice you have to check if coverage, obtained with distribution of computers device to check, allows more than one path towards Master device. In case of paths supported only by sequence of one device at a time, the loss of an intermediate element blocks the communications, because of the lack of a possible alternative path.
 - *Turn all devices provided for network on and try to simulate the shutdown of some intermediate Routers, checking on Master the right reception of messages also from the farthest devices.*
6. Add Routers in case you want increase the redundancy of paths in order to give the network more than one transmission possibility.
7. Configure the sleep time of devices in order to obtain an optimal configuration with low energetic consumption mode.
 - *In 3DOM for all the Routers and Coordinator, set the sleep time so that it is higher (almost twice) of the longer time programmed in any networked Slave; in order to obtain a good functioning of the network, is important that all Routers and Coordinator to share the same value; this value is irrelevant for the End Device.*

The above described principle for network setup can be improved during first network start-up, as the instruments, equipped with display, will allow the identifying of all elements that can be reach through RF and the evaluation of connection quality of signal displaying the diagnostic of device (Index SS = value from 0 to 100%; see §4.3.4). In any case an estimation of physical layout of network is useful in order to foresee the needs of further support items. Network isn't absolutely binding from point of view of present components, so you can modify the structure, through Master, any time you need it.

Note 1: in a network, in case of reprogramming of Master without modify of configuration of Slave instruments (which are turned on and active in acquisition mode), it could happen that the Slave devices are not connected to the network and data from these stations don't arrive (or arrive at different delay times). In order to restore the correct operating mode, restart the Slave devices or set the fast acquisition mode of Slave to have available quickly a great number of connection attempts to the network.

Note 2: on Slave devices it's always possible turn on manually the radio by pressing of  button from diagnostic mask 7 (the mask dedicated to ZigBee Radio). There is also a configuration parameter (programmable from 3DOM software) that determines the fixed power of radio for Slave

devices (if you want to interrogate them at any time); in this case, these are supplied by means of external power supply pack (consumption about 20 mA continuous).

5.6.3. Ethernet

R-Log can communicate through the use of Ethernet LAN/WAN network, so virtually of any length. R-Log offers to the user two possibilities in order to use this characteristic:

- external RS232/Ethernet converter connected to acquirer through ELA105 cable, supplied together with R-Log; sale code LSI LASTEM: DEA550; see §5.6.3.1 for its configuration;
- Ethernet converter built-in datalogger, available only for R-Log model ELR516M; see §5.6.3.2 for its configuration.

5.6.3.1. DEA550 converter configuration

In order to configure RS232/Ethernet converter prepare a PC with *HyperTerminal* of Windows and follow the instruction below:


1. Connect the converter to its power supply pack and to PC through proper serial cable (supplied together with DEA550) like specified in sheet supplied with product. Do not turn device on.
2. Set-up the switches for selection of serial type on “RS232” and red switch, placed opposite to serial, in “console” position; see plate on back of converter in order to set-up the position of switches.
3. At HyperTerminal opening, after you have given the name to work session, in “configure” session, set-up the exact COM to which converter is connected to PC; finally configure HyperTerminal with *9600, N, 8, 1, Nessuno parameters* referred to port where DEA550 has been connected.
4. Turn converter on and await for inscription “Login” on program. After login has been executed, the symbol # will be displayed; so type in “root” and then password “root” (press *Enter* key to continue the procedure an *Esc* key to come back).
5. Type in *EDITCONF* and press *Enter* key.
6. In order to change TCP/IP address, from main screen, select 1(Network configuration) and then again 1(IP configuration). Now modify values 2(IP address) and 3(Subnet mask) with values supplied by user.
7. Default port is 6000. In order to modify it (*Port*), from main screen, select 2(Serial port configuration), 1, 3(Host mode configuration) and finally 2(Port number) and here modify the value. Recommended port is 7001.
8. Change *Inter character time-out* from 0 to 10 ms. From main screen, select 2(Serial port configuration), 1, 4(Serial port parameters) and finally 7(Inter character time-out).
9. Type in sequence (always followed by *Enter*), *Save*, *Apply*, *Esc*.
10. Turn converter off and put red switch in position *data*.
11. Finally turn converter on again and connect it to R-Log.

5.6.3.2. XPORT configuration (internal Ethernet port)


For right operation, each R-Log must be configured according to special setups (for TCP/IP protocol) through DeviceInstaller of Lantronics software, available in DVD LSI LASTEM products (MW6501) in “Install\Support\Lantronix” folder or downloadable from Lantronics website <http://www.lantronix.com/support/downloads.html>

To configure Ethernet port, follow the instruction below:

1. Connect LAN cable in proper port on top panel of R-Log and check the flashing LED on the port, indicating the presence of the network.
2. Open DeviceInstaller software of Lantronics and following window will be displayed:




3. Press *Search* and identify the computer device connected to LAN. After some seconds, if LAN connection is operating properly, the instrument is recognized and listed among the available ones. In case of several instruments, it's possible to determine each model according to hexadecimal numeric code written on label placed on bottom side of box.




4. The instrument use default DHCP protocol to obtain IP network address. Fix a static IP address and other indispensable network parameters. Ask these information to your administrator; in any case note these setups as you'll have to enter them into communications setups of 3DOM. To modify the operation parameters: double click on device on right side of window, then select *Web configuration* session and click on green arrow button. Login window will be displayed asking for user name and proper password:


like default the instrument is supplied with null user and password, so confirm immediately the not compiled input fields. Mask below will be displayed:



5. Select *Network* session for programming of network parameters. Set the parameters previously selected (remember that the default setting of port is 10001, then, if possible, set that port in *Network->Connection->End point configuration->Local port*; if it isn't possible, choose another port and set this in 3DOM program too) and then click on *Apply settings* command; await for some instants and repeat the starting operations to carry out new recognition of instrument, that now is operating in LAN network with new set parameters.



6. Select *Serial settings* session to modify setups of serial communication port. The only one possible modification is *Baud rate*. Choose and Select values from 1200 to 115200 bps, making attention to specify, through 3DOM program, same value for serial port 1 of LSI LASTEM instrument.



7. For modification of other advanced configurations make reference to online help of *DeviceInstaller* program.


5.6.4. Data trasmission through GPRS connection

(function available only for ELR515 and ELR516 serie models)

GPRS data transmission system consists of the following parts:

- One or more R-Log instruments;
- One modem GPRS LSI LASTEM Mod. DEA717 or DEA718 for each R-log, connected through ELA110 cable; the modem SIM card must be able to GPRS data transmission and the request of PIN code must be disable;
- One PC server with Windows operating system (Window XP or next version, Windows Server 2003 or next version), connected to Internet with a public IP address;
- LSI LASTEM *CommNetEG code BSZ306.2* program with use licence enabled for GPRS connections (only for transmission through TCP socket).

R-Log must be configured through *3DOM* program, in order to use GPRS modem and transfer the elaborated data with selected timing; the GPRS modem can be used only with serial port 1. In the configuration through *3DOM* software, remember to enable with *Yes* the field *Modem powered with actuator* in the *Serial communication port 1* window when the modem is powered using actuator 3 (recommended solution).

GPRS connection happens according to the spontaneous transmission rate programmed inside the system; therefore it happens according to the programmed timing (on the initiative of the instrument); you can also make a manual data transfer by pressing the button . The remote computer (that collects the data) consists of one TCP server, set on IP public address. The address of this server has been programmed inside GPRS modem connected to R-Log.

There are two different transmission's modes:

- Through TCP socket;

- Through FTP protocol.

The TCP socket connection allows the use of minimum quantity of IP packets, and for this reason it's cheaper than FTP protocol. Moreover one server (no special software installed in it) can transfer the data through TCP protocol.

In addition to the above-described parameters, set the following parameters through 3DOM software in according to the selected transmission's mode:

- for TCP mode, select *Switched off and on in case of failure* option in *Serial communication port 1* window;
- for FTP mode, select *Power on/off due to actuation logic* option in *Serial communication port 1* window to set the modem to switch off once a day. Then configure an actuation logic for leave the modem turned on all day except for a minute that will remain turned off; it is advisable to turn on the modem at least two minutes before the transmission.

R-Log will transfer the data to the operative centre at set timing. The parameters of the operative centre (IP address, IP port, user name, etc) have been specified inside *config.ini* configuration file (stored inside GPRS modem). In case of transmission through TCP socket the operative centre must be equipped with LSI LASTEM *CommNetEG code BSZ306.2* program. In case of transmission through FTP protocol, one FTP server must be available and working.

These are the reports of operational errors in GPRS transmission:

Symptom	Setting
Data not received through FTP/TCP socket	First installation

Cause-reason	GPRS diagnostic window	Resolution
Modem installation not correct	FSM: 3 and then 3 times GPRS Modem: N.C.	Check the connections of the serial cable and power supply (either directly or with actuator) according to the documentation.
Modem isn't configured for GPRS connection	FSM: 3 and then 3 times GPRS Modem: N.C.	Contact LSI-Lastem.
SIM protected with <i>pin</i> code or not present or not suitable	FSM: 3 and then 3 times GPRS modem: N.C.	PIN code request must be disabled (use a normal cellular phone to do it). The SIM must be enabled to receive and transmit GPRS data.
COM 1 Port of R-Log not appropriately configured	FSM: 3 and then 3 times GPRS Modem: N.C. or no GPRS window	With 3DOM check the COM 1 port settings on the current instrument configuration, in particular the type of modem, power supply, the modem power type, the protocol type and the rate of spontaneous GPRS transmission.
Elaborations not available on R-Log (usually happens when you start a transmission in manual mode by pressing the suitable button on the instrument)	FSM: 11	Wait R-Log elaborations. It's advisable to set the rate of spontaneous GPRS transmission greater or equal to the elaboration rate.
Low CSQ signal (<16)	FSM: 3, 7, 10 CSQ<16	Try to position the modem so that the signal is higher; possibly replace the antenna. CSQ values: 0÷32 (15 poor, 32 excellent).
Wrong FTP site or TCP address in modem configuration	All regular	Edit the configuration file of the modem. For more informations refer to the manual provided with the modem.

Symptom	Setting
Data periodically received through FTP/TCP socket	Installation has already occurred

Cause-reason	GPRS diagnostic window	Resolution
Low CSQ signal (<16)	FSM: 3, 7, 10 CSQ<16	Try to position the modem so that the signal is higher; possibly replace the antenna. CSQ values: 0÷32 (15 poor, 32 excellent).
The modem is turned off		If the R-Log doesn't turn off, the issue depends by actuator configuration of the modem (actuator #7). Reprogram the actuator. If E-Log turns off, see the cause "E-Log is turned off".
R-Log is turned off		Could be a R-Log power problem. In order to know when the last reset is happens, go to the instrument diagnostic window type 2 (§4.3.3 R-Log manual).

5.6.4.1. Use of DEA717 GPRS Industrial Plus modem

In order to install and configure DEA717 modem, see INSTUM_00519 manual supplied with the modem.

5.6.4.2. Use of DEA718 GT863-PY Telit modem

In order to install and configure DEA718 modem, see INSTUM_00518 manual supplied with the modem.

5.7. Operating of low energetic consumption

There isn't a very configuration for low energetic consumption mode. Anyway in order to optimize energy consumption, proceed as follows:

- Use the highest possible value to program the acquisition rate of the measures, according to the dynamic feature of the quantity's signal to be sampled (see §5.2);
- Keep the actuation time to energize the sensors low without compromising the reliability of the measure;
- Use the lowest possible acquisition time required to, however, have the sensor send the measured signal in a correct manner;
- Use a high rate, and only if necessary, to program the rate for probe control (see §5.2.6);
- Switch the display off manually or program it to automatically shut off by itself (see §4.3.5);

the display can be turned off by pressing of  button, while displaying the mask of measures;

- Disable the fast acquisition mode of the measures which was possibly enabled while the instrument was switched on (see §5.2.5).
- Remove not used actuation logics (see §5.5);
- Manage modem switch on through actuator 3 (programmed with timed logic), or programming GPRS communication through modem activation;
- In case of sensors or telephone modems fed continuously, don't use instrument's batteries, but use other ones (if possible). In this way the instrument can operate even if modem and/or sensors aren't fed, and so they aren't operating;

ATTENTION: Default configuration programmed on instrument is not the one of lowest energetic consumption.

In case of Slave devices:

- The Slave devices must have a channel to measure battery (level battery or voltage) that bypasses the internal acquisition mode, normally set at one minute. The low energetic consumption is achieved by programming the measure with a higher acquisition rate (5 minutes or more recommended).
- It's important don't turn off (accidentally or otherwise programmed) the device "parent" of the Slave because before the next attempt to send data, the Slave should enter into a search mode (which still has a limited duration) that consumes a lot of energy.

6. Appendixes

6.1. Technical specifications

A/D Convertitor	Resolution: 18 bit rounded to 16 bit Sample duration (rejection 50/60 Hz): 80 ms @ rejection 50 Hz		
Analogue inputs	Number of analog inputs: 4 in differential mode, 8 in single-ended mode		
		Scale	Resolution
	mV	-300 ÷ 1200 mV	40 µV
		±78 mV	3 µV
		±39 mV	1.5 µV
	Pt100	-50 ÷ 125 °C	0.003 °C
		-50 ÷ 600 °C	0.013 °C
	Resistance Ω	80 ÷ 140 Ω	0.0013 Ω
		80 ÷ 320 Ω	0.005 Ω
		0 ÷ 6000 Ω	0.19 Ω
	Thermo couples	E-IPTS 68 -200 ÷ 1000 °C	< 0.1 °C
		J-IPTS 68 -50 ÷ 600 °C	< 0.1 °C
		J – DIN -50 ÷ 600 °C	< 0.1 °C
		K-IPTS 68 -150 ÷ 1350 °C	< 0.1 °C
		S-IPTS 68 0 ÷ 1600 °C	< 0.22 °C
		T-IPTS 68 -200 ÷ 200 °C	< 0.1 °C
	ESD protection : <ul style="list-style-type: none"> ±8 kV contact discharge IEC 1000–4-2 ± 1.2 kV air-gap discharge IEC 1000–4-2 		
	Channel to channel crosstalk: -93 dB Max input signal: 3 V Every input with EMC filter		
	Error as a function of Environmental Temperature: <ul style="list-style-type: none"> scale -300 ÷ 1200 mV < ±0.01% ES (@ -10 ÷ 30 °C) scale ±39 mV < ±0.01% ES (@ -10 ÷ 30 °C) scale ±78 mV < ±0.01% ES (@ -10 ÷ 30 °C) (ES= end of scale) Cold joint error as a function of Environmental Temperature: <ul style="list-style-type: none"> Thermocouple E-IPTS 68 ±0.029/°C (@ 0 ÷ 50 °C) Thermocouple J-IPTS 68 ±0.02/°C (@ 0 ÷ 50 °C) Thermocouple J-DIN ±0.017/°C (@ 0 ÷ 50 °C) Thermocouple K-IPTS 68 ±0.025/°C (@ 0 ÷ 50 °C) Thermocouple S-IPTS 68 ±0.05/°C (@ 0 ÷ 50 °C) Thermocouple T-IPTS 68 ±0.038/°C (@ 0 ÷ 50 °C) 		
Digital input	Functions: <ul style="list-style-type: none"> Input for sensors with optoelectronics (freq. max 5 kHz); Frequency input (freq. max 5 kHz); Logic state input ON/OFF (they acquire signals 0 ÷ 3 Vdc). State level “low”: 0 ÷ 1.5 V State level “high”: 2 ÷ 3 V Max input frequency: 5 kHz Max error.: 3 Hz @ 5 kHz Protection: Transient voltage suppressor: 600W, t > 10 µs		
Total measures	50 (total measures referred to sensors acquired from terminal board or serial port/radio and calculated measures)		

Derived calculations	<p>Function library for derived calculations applied to acquired measures or calculated:</p> <ul style="list-style-type: none"> • Mobile statistic (minimum, arithmetic average and ed angular, maximum: statistic sliding from 2 to 128 values) • Crossed statistic (average and instantaneous totalization from 2 to 5 measures)
Actuators output	<p>Nr. outputs: 2 (+1 on pin9 of the connector of serial port 1)</p> <p>Max current for a single powered output: 150 mA</p> <p>Protections: thermal and for overloaded current (> 0.15 A)</p>
Power	<p>Input voltage: 8 ÷ 14 Vdc charger</p> <p>Average power consumption (without probes feed, @ 4.2 Vdc):</p> <ul style="list-style-type: none"> • Illuminated display power on: 200 mW • 4 channels acquisition: 2 mW • Radio power on: 200 mW • Power down: < 1 mW <p>Protections:</p> <ul style="list-style-type: none"> • Internal overcurrent, shortcircuit • Max actuator current, with external load: 1.5 A • Transient voltage suppressor: 600 W, t > 10 µs • Polarity inversion
Internal battery	<p>Internal battery: rechargeable LiIon 1950 mAh, 4.2 Vdc</p> <p>Recharge time: about 8 hours</p> <p>Discharge time (energetic autonomy exclusively with a new and full charged battery): about 9 months</p>
Actuation logics	<p>3 actuators with 10 actuation programmable logics; activation of outputs according to AND o OR logics applied to one or more actuation logics:</p> <ul style="list-style-type: none"> • Feeding of sensors with automatic anticipation • Daily timer (hour of activation and hour of deactivation) • Cyclic timer (ON status time, OFF status time) • Overcoming of threshold value (limits and hysteresis) • Logics for outdoor applications (according to the R-Log model): eolic alarm, evaporimeter filling, flood alarm, snow level alarm) • Manual power on and power off
Measures elaborations	<p>Statistical algorithms for arithmetic and vectorial elaborations, uniquely programmable for each measure, calculated on single time base, from 1 second to 12 hours, for all measures. (default elaboration rate of measures: 10 minutes)</p>
Communication protocols	<p>Serial port: Native (CISS) for instrument configuration, internal clock date/time setting, data download (instantaneous and elaborated measure values and diagnostic informations); GPRS transmission support (TCP or FTP mode).</p> <p>Radio: data acquisition from probes or LSI LASTEM datalogger with CISS protocol</p>
Communication lines	<ul style="list-style-type: none"> • RS232 main serial line, communication speed from 1200 to 115200 bps, function of instrument configuration and reception of instantaneous and elaborated data through LSI LASTEM protocol; Bluetooth connection option • Ethernet RJ45 connection with TCP/IP protocol • RS232 second serial line, communication speed from 1200 to 115200 bps; function of instrument configuration and reception of instantaneous and elaborated data through LSI LASTEM protocol; availability of MODBUS and others protocols; connection to GSM/GPRS modem
Radio	<p>Type: ZigBee</p> <p>Frequency: ISM 2.4 GHz, 13 direct sequence channels</p> <p>Diffused power: 10 mW (+10 dBm)</p> <p>Reception sensivity: -102 dBm</p> <p>Standard range: 60 m for indoor applications, 500 m for outdoor applications (as the crow flies)</p> <p>Function: transmission and reception of instantaneous and elaborated data</p>
Internal calendar clock	<p>Format: year-month-day-hour-minutes-seconds</p> <p>Drift: 30 seconds/month (with T=25°C)</p>
Display	<p>Alphanumeric, 4 x 20 charaters, illuminated with programmable power off</p> <p>Functions displayed:</p> <ul style="list-style-type: none"> • Instantaneous values of acquired sensors and calculated measures; • Signal of measures in allarm according to actuation logics; • Diagnostic of actuators state; opportunity of instantaneous modify of actuator activation state;

	<ul style="list-style-type: none">• Diagnostic of communication lines and protocols;• Communication diagnostic (radio and GPRS);• Diagnostic data memory;• Product presentation window of the instrument.
Keyboard	Membrane, 8 keys, indicators (LED) that display the operation state of instrument (OK/ERR), the communication from and to external devices (Rx/Tx) and the charge-condition of inner battery
Processors	1 RISC 8 bit, clock 8 MHz
Mechanic protection	IP 40
Operating temperature	-20 ÷ 60 °C, 15 ÷ 100 % UR (no condensing)
Weight	About 500 g
Dimensions	140 x 120 x 50 mm

6.2. Function library for derived calculations

R-Log has a useful library containing deriving quantities, featuring application functions dedicated to both indoor (microclimate) and outdoor (meteorology) environmental sectors.

The list below shows all calculation functions available:

- **Arithmetical calculations**
 - Addition-summation
 - Subtraction
 - Multiplication
 - Division
- **Statistical/mathematical operations**
 - Integral
 - Mean
 - Power
 - Exponential
 - Natural logarithm and base 10
 - Square root
- **Mobile calculations (programmable from 2 to 128 values)**
 - Minimum, average, maximum
 - Totalization
 - Angular average (only for ELR515 and ELR516 models)
- **Thermo-hygrometric quantities (UNI EN ISO 7726, ISO/WD 7730, VDI 3786)**
 - Relative humidity with psychrometric calculation (dry/humid bulb)
 - Absolute humidity
 - Specific humidity
 - Mixing factor (ratio)
 - Humid air enthalpy
 - Dew point temperature
 - Partial steam pressure
 - Humidity index (HI)
 - Discomfort of heat index
 - Indoor and outdoor WBGT index
 - Wind chill index
 - Chill temperature (TCH)
 - Mean radiant temperature
 - Radiant temperature asymmetry (only for ELR510 models)
 - Average planar radiant temperature (only for ELR510 models)
 - Planar temperature side 1 and side 2 (only for ELR510 models)
 - Percentage of dissatisfied people due to radiant temperature asymmetry from wall or ceiling (only for ELR510 models)
 - Dissatisfied floor temperature (only for ELR510 models)
 - Dissatisfied vertical temperature from 10 to 110 cm (only for ELR510 models)
 - Draught rating (only for ELR510 models)
 - Operative temperature (only for ELR510 models)

➤ **Duct flow**

- Air speed from differential pressure (Pitot or Darcy)
- Volumetric air and mass flow
- Number of air changes

➤ **Radiometry and illuminance**

- Insolation time
- UV index (DLE)
- UV exposure level
- Light intensity (only for ELR510 models)
- UVA density (only for ELR510 models)
- Day light factor (only for ELR510 models)

➤ **Actuators operations(*)**

- Actuation status calculated with AND operator
- Actuation status calculated with OR operator
- Actuation logics status calculated with AND operator
- Actuation logics status calculated with OR operator

➤ **Others**

- Evaporation calculation based on the evaporimeter level (only for ELR515 and ELR516 models)
- Soil volumic humidity with permittivity
- Atmospheric pressure at sea level
- Total count
- Delta with previous value
- Detection and warning of low battery status of one or more sensors received through CISS protocol

(*) Calculated measures available only for R-Log datalogger with firmware from 2.3 version; calculated measures associated with 3DOM software from 3.8 version.

These measure are set a fixed update time of 1 second: the update of value of calculated measure on the actuator is always 1 second delay compared to the status recorded by internal actuation logic.

Now is therefore possible to process and record the actuator status or display it as an instantaneous value without having to carry the electrical signal of the actuator in an input of the datalogger. Now it is also possible make complex algorithms through a combination of actuation logics status with other calculated measures (typically addition and multiplication), overcoming the limits of the use of a single operator (AND and OR) and giving the inputs for the acquisition of real sensors.

With *3DOM* you can choose which quantities to have the instrument calculate and select the direct measures which allow their calculation.

Some calculations, frequently used in agro-meteorology, such as wind direction or integral radiation, are easily obtained from the integral mathematical calculation.

6.3. Error messages

R-Log indicates errors by triggering the red led *Err* located on the instrument's keyboard: the kinds of errors, listed in the following table, are indicated by the blinking modes.

Number of blinkings	Type of problem	Troubleshooting
1	Data memory access	Try to transfer the data elaborated by the instrument; then send the configuration again to the instrument using <i>3DOM</i> ; if the error occurs again, contact LSI LASTEM technical support
3	Acquisition from sensors	Use <i>3DOM</i> to check the congruence of the acquisition parameters of the measures in the current configuration; send the configuration again to the instrument; should a new error occur, contact LSI LASTEM technical support for assistance including the current configuration file
4	Parameter configuration	Transfer the data elaborated by the instrument; then send the configuration again to the instrument using <i>3DOM</i> ; if the error occurs again, contact LSI LASTEM technical support for assistance including the current configuration file
5	Search data in memory	If the error occurs immediately after switching on the instrument, reset the error and continue using the instrument as usual; should a new error occur while operating the instrument with no communication with the PC, contact LSI LASTEM technical support; should the error occur while requesting the elaboration in memory by the program <i>3DOM</i> , try requesting the data again using a different date/time (try using a previous date first; should a new error occur, try later dates); should the error persist, switch the instrument on and off and request all data indicating a date/time definitely prior to the first date in memory; upon completion of the data transfer, cancel them; should the error occur again send the configuration of the instrument to LSI LASTEM technical support describing the problem.
6	Protocol CISS	Use <i>3DOM</i> to check the congruence of contents parameters in the current configuration; send the configuration again to the instrument; should a new error occur, contact LSI LASTEM technical support for assistance including the current configuration file

When the instrument finds an error, it also indicates it as follows:



- 1) By activating the digital signal of the error;
- 2) By displaying the numeric value of the error found: the error code appears on the visualization window of the diagnostic data type 1 (see §4.3.4); the numeric code can actually indicate more than one error; the displayed value is expressed with a hexadecimal

number; the errors that correspond to the value that appears on the display can be easily interpreted by means of program *3DOM* (menu *Instruments*);

- 3) By communicating with program *3DOM*: while the instrument transfers the elaborations or sends the configuration data, the programs may show the code of the error found by the instrument.

6.3.1. Disabling error indication

There are two possibilities to disable error indication:

- 1) Press key  and  while the error code is being displayed in the diagnostic window type 1 to reset the error (it will however remain on the display when you leave and call the diagnostic window);
- 2) While the data is being transferred between instrument and PC: in this case the instrument resets the error since it considers that the error has already been found by the person using the computer.

In both cases the reset operation will locally switch off the optical indicator *Err*, eliminate the error number from the diagnostic window type 1 and disable the signal in effector's output (if programmed by respective actuation logic). The error reset will remain until the instrument finds a new possible error; in this case the instrument will activate the error indication mode again as previously described.

6.3.2. Error found in measure

The values indicated by the measures (instantaneous values) may indicate the error status (*Err*), in the following conditions:

- If the measure is acquired:
 - Programming of a measure incompatible with the type of signal to be measured (electrical type selected, linearization type, scale recalculation parameter, etc.);
 - Sensor is interrupted or is not correctly connected to the terminal board;
 - If the sensor is powered by an energized output, the starting time might not be enough;
 - The electrical input signal (to the instrument) is out of scale;
 - In case of thermocouple measure, the cold junction temperature (internal temperature) might not be programmed;
 - In case of measure acquired from serial port: no valid message has been detected coming from the sensor during time corresponding to more then treble set acquisition rate.
- If the measure is calculated:
 - Out-of-scale or in-error value from one or more dependant measures allocated to the measure to be calculated;
 - Error in output to the algorithm of the calculated measure;

When the number to be displayed is < -999999 or > 9999999 , E-Log points out the anomaly of measure in overflow with "Overfl." error status.

6.3.3. Low battery alarm



This condition can occur only for Slave R-Log instrument because Master and Repeater instruments should be always powered by external power supply.

The low battery alarm occurs on Slave datalogger through three times blinking of LED ■-Err; the instrument also sends a low battery signal to the Master datalogger for every spontaneous data transmission. The error LED blinking disappears automatically in Slave instrument when the battery becomes charged.

The Master instrument detects the low battery alarm comes from the Slave instrument and it also flashes the LED ■-Err three times; it signals the “20000000” error on the diagnostic window type 1 and for the first instantaneous value related to the sensor in error, it signals the “LB” indication on

the window of instantaneous values (press twice the  button to display it).

Despite the error condition disappears automatically at the return of the correct level of power supply, the Master instrument however signals persistently the error on the system window; in order

to stop the error signal press at the same time  and  buttons on diagnostic window type 1, then exit and reentry in the same window; the error signal disappears automatically in window of instantaneous value.

In the following table there are values of battery voltage to trigger the alarm condition:

<i>Condition</i>	<i>Battery voltage</i>
Minimum voltage of battery to “keep alive” the instrument	3,3 V
Minimum voltage of battery to trigger the alarm condition	3,5 V
Minimum voltage of battery to operate with the survey	3,7 V

6.4. Instrument maintenance

R-Log does not require special maintenance interventions when installed according to the general safety rules indicated in §1.

However, we recommend that LSI LASTEM staff carries out periodical check of whole installation (E-Log and sensors connected to it) at regular intervals in order to point out and correct possible measurement errors.

6.5. Declaration of conformity



LSI LASTEM S.r.l.
Via Ex S.P. 161 Dosso, n.9 - 20090 Settala Premenugo (MI) - Italia
Tel.: (+39) 02 95 41 41 **WEB:** <http://www.lsi-lastem.it>
Fax: (+39) 02 95 77 05 94 **CF./P. Iva:** (VAT) IT-04407090150
e-mail: info@lsi-lastem.it **REA:** 1009921 **Reg.Imprese:** 04407090150



DICHIARAZIONE DI CONFORMITA' CE

Declaration of Conformity

Produttore: **LSI LASTEM s.r.l.**
Applicant: Via Ex S.P. 161 Dosso, n.9 – 20090 Settala Premenugo (MI) – Italia

Con la presente si dichiara che tutti i prodotti della seguente serie:
We hereby declare that all the products of the following series:

Datalogger per applicazioni ambientali modelli:
Datalogger for environmental applications models:

- **ELR510 – ELR515 – ELR516**

a cui questa dichiarazione si riferisce è conforme ai requisiti essenziali dei seguenti standard e documenti normativi:

to which this declaration relates, is in conformity with the relevant provisions of the following standard and other normative documents:

ETSI EN 300 328 v 1.7.1 (2006-10)
ETSI EN 301 489-1 v 1.8.1 (2008) - ETSI EN 301 489-3 v 1.4.1 (2002)
ETSI EN 301 487-17 v 2.1.1 (2009-05)
EN 62311 (2008)
EN 61000-4-2 (1995) + A1 (1998) + A2 (2001)
EN61000-4-3 (2006) + A1 (2008)
EN61000-4-4 (2004) - EN 61000-4-5 (2006) - EN61000-4-6 (2007)

che rispettano le direttive:
following the provisions of the Directive:

R&TTE Directive (1999/5/EC)

Settala, giugno 2011

Dr. Giulio Certo
Direttore Generale e Legale Rappresentante

CE_00734_it-en_0



Thank you for reading this data sheet.

For pricing or for further information, please contact us at our UK Office, using the details below.



UK Office

Keison Products,

P.O. Box 2124, Chelmsford, Essex, CM1 3UP, England.

Tel: +44 (0)330 088 0560

Fax: +44 (0)1245 808399

Email: sales@keison.co.uk

Please note - Product designs and specifications are subject to change without notice. The user is responsible for determining the suitability of this product.