



CellarSafe

Carbon Dioxide and Oxygen Safety Monitor



**Installation and operating instructions
for the CellarSafe Monitor
and the CellarSafe Calibration Unit**

M07234
Issue 5 05/2006

1 INTRODUCTION

1.1 Product overview

The CellarSafe is a carbon dioxide and oxygen fixed monitoring system designed to monitor gas levels in a confined space, such as a beer cellar. The CellarSafe consists of one base unit housing the gas monitoring system and is supplied with one repeater unit to be located in a safe position outside of the confined area. The CellarSafe is supplied as a carbon dioxide monitor only, or as a combined system with the addition of an oxygen monitoring system. Both formats have two alarms; a pre alarm and a main alarm. The carbon dioxide monitoring system looks for rising gas levels. The carbon dioxide and oxygen combined monitoring system monitors rising gas levels for carbon dioxide or falling gas levels for oxygen. The CellarSafe operates two alarm relays if the gas level exceeds the two pre-set alarm levels. The output from these relays can be used to control external visual or audible alarms, or ancillary equipment such as a fan.

The CellarSafe is designed for easy installation using mains power supply. It is pre-calibrated and has been designed for one man installation. The alarm levels are factory set and the relays are 'fail safe' so in event of power failure the alarms will activate.

The CellarSafe is supplied with one repeater unit. The repeater unit provides a remote interface to mimic the gas monitoring alarms, fault alarm, power indicator, MODE button and sounder of the base unit. A series of repeater units can be daisy chained together to provide an extension of monitoring system. Each additional repeater unit is provided with a 9 m cable.

1.2 Product description

The CellarSafe consists of two components; a base unit and a repeater unit. The base unit is a single shell housing the gas monitors: a front operator display panel containing a liquid crystal display, MODE button, four illuminating LEDs and a sounder (see "Figure 1.1"). The CellarSafe is provided with a 2 m mains cable, two carbon dioxide warning labels, one repeater unit (see "Figure 1.2") with a 9 m connection cable. Electrical connections for mains power, two relay outputs and the repeater unit are located at the bottom of the base unit. The repeater unit mimics the base unit but does not have the display panel.

The illuminating LEDs on the base unit and all repeater units are as follows:

- 'POWER', green LED
When flashing indicates the system is running. If this LED is extinguished the system is not running, if the LED is on continuously a fault condition is present.
If the battery backup option is fitted, the sounder will beep every 20 seconds in the event of mains power failure.
- 'FAULT', yellow LED
When flashing indicates there is a system fault.
- 'LOW ALARM', red LED
When flashing indicates the low alarm level (pre alarm) is exceeded.
- 'HIGH ALARM', red LED
When flashing indicates high alarm level (main alarm) is exceeded.

When the low or high alarm gas level is exceeded the sounder and alarm LEDs will operate, "Table 1.1:" outlines the alarm conditions.

Table 1.1: Alarm conditions for the CellarSafe and repeater units

Gas levels	RELAY INDICATION	LED INDICATION	SOUNDER INDICATION
Normal	Not active	No alarm LEDs	No sound
Increasing gas levels of carbon dioxide			
Gas level > Low Alarm	Low Alarm	Low Alarm LED flashing	Long beep once per second
Gas level > High Alarm	Low and High Alarm	Low and High Alarm LEDs flashing	Long beep once per second

Gas levels	RELAY INDICATION	LED INDICATION	SOUNDER INDICATION
Decreasing gas levels of oxygen			
Gas level > Low Alarm	Low Alarm	Low and High Alarm LED flashing	Long beep once per second followed by one short beep
Gas level > High Alarm	Low and High Alarm	Low and High Alarm LEDs flashing	Long beep once per second followed by one short beep

The MODE button is used to communicate with the CellarSafe after an alarm is triggered. In the first instance pressing and holding the MODE button for 5 seconds will silence the CellarSafe alarm sounder, the alarm LEDs will continue to flash in an alarm condition (i.e. gas is still present or in the case of oxygen, depleted). When the gas level has returned to normal the illuminating alarm LEDs will extinguish but will reactivate the sounder and illuminating LEDs if the alarm level is exceeded again. The CellarSafe is set to use non-latching alarms the LEDs will automatically extinguish when the gas level returns to normal.

If the system fault alarm is activated the CellarSafe will sound a beep twice every second and the yellow warning light will flash. The system fault alarm will activate if a sensor fault condition occurs.

Electrical connections for external alarm equipment are located within the main CellarSafe base unit. One connection terminal is provided for the main alarm and one for the pre alarm.

Figure 1.1 The CellarSafe with carbon dioxide and optional oxygen monitors.





Figure 1.2 The CellarSafe Repeater Unit.



1.3 Options for the CellarSafe

The CellarSafe can be fitted with the addition of an oxygen monitor to provide dual monitoring of safe gas levels of carbon dioxide and oxygen. The CellarSafe can be fitted with a backup battery providing 2 hours of backup in the event of mains power failure. When in backup mode the CellarSafe will emit a beep once every 20 seconds. Up to 4 repeater units can be used with the CellarSafe base unit. Each additional repeater unit is supplied with a 9 m cable. For a list of CellarSafe units and options see "Table 1.2".

Table 1.2: CellarSafe units and options

ITEM	PART NUMBER
CellarSafe - complete unit	Contact Crowcon
CellarSafe with battery backup	Contact Crowcon
CellarSafe with oxygen sensor module	Contact Crowcon
CellarSafe with battery backup and oxygen sensor module	Contact Crowcon
† Repeater unit including cable	99S38001
115 V Model	Available on request
Field Calibration Unit	99S38002

†Additional repeater units come with 9 m cable.

Important

The battery backup unit and the oxygen option contain non-environmentally friendly elements. Please dispose in an environmentally and safe manner or return to Crowcon for safe disposal.

2 INSTALLATION

When carrying out any installation work, ensure that local regulations and site procedures are followed.

2.1 General

This manual contains instructions for installing the CellarSafe base unit and repeater unit. Before beginning installation of the CellarSafe please follow the mounting, cabling and installation instructions paying close attention to the location of the base unit.


2.2 Mounting

The CellarSafe base unit is designed for wall mounting and should be located near to the valves and manifold within the confined space approximately 0.5 m from ground level. Fixing holes are attached to the unit so that it can be attached to the wall with appropriate screws and wall plugs.

The CellarSafe repeater unit should be mounted outside of the confined space, for example, the doorframe at the entrance, using the fixing brackets provided. The height should be at eye level for ease of use. Additional repeater units are mounted in the same way, suitable locations are outside of the confined area, for example, at a secondary exit or positioned at a different location normally occupied by staff, in a safe position, within the building.

The CellarSafe is not suitable for outdoor or hazardous environments.

Figure 2.1 example location and mounting



2.3 Cabling requirements

CellarSafe is fitted with an AC plug for convenience, however if permanent installation is required, it should be wired to a dedicated AC spur fed from a circuit breaker, marked as the disconnecting device for the CellarSafe.

The circuit breaker must comply with the relevant requirements of IEC 947-1 and IEC 947-3. These procedures ensure that the installation complies with the requirements EN 61010-1 (Low Voltage Directive).

The repeater unit supplied with CellarSafe does not need specialist cabling. Each additional repeater unit is supplied with a 9m cable as standard.

2.4 Electrical connections for external alarm equipment

The CellarSafe has two alarm relays to which external visual and audible alarms or other ancillary equipment, such as a fan, can be connected. An example electrical connection is shown in "Figure 2.2". Before connecting the equipment, ensure the mains supply is switched off. Unscrew the top cover to access the screw terminal connectors on the PCB board. The two alarm relay connections are marked NC, NO and C. Use the NO and C screw terminal connectors as shown in "Figure 2.2".

NOTE: The alarm relays are shown in the normally energised state. When installing the device the normally open (NO) connection will be closed since there is no power. Similarly, the normally closed (NC) connection will be open.


The CellarSafe can be connected to two fans, one to each alarm relay. If the pre alarm is exceeded the first fan will activate. If carbon dioxide levels increase further and the main alarm is then exceeded,

the second fan will activate to increase ventilation. Dual-speed fans are also available for a similar effect: low speed and high speed corresponding to low and high alarm respectively.

It is recommended that external equipment should be fused.

NOTE: The relay contacts must not be used to switch loads drawing more than 10 Amps.

Figure 2.2 Alarm relay electrical connections.



3 OPERATION AND SET UP

3.1 Mode button

The CellarSafe base unit and repeater unit provide a MODE button to communicate with the CellarSafe. The MODE button is used in the event of an alarm or fault occurring, and during the self testing procedure.

3.2 Operator display panel

The operator display panel provides a reading of the concentration of gas levels being monitored. Carbon dioxide, CO₂, is shown as a percentage concentration level and at normal gas level (i.e. safe levels) will display 0 to 0.2%.

With the oxygen monitor option, the display will alternate between showing the carbon dioxide gas concentration, CO₂, and the oxygen gas level concentration, O₂. Normal oxygen gas level concentration is 20.9%.

To activate the display backlight, press the MODE button.

3.3 Set up

The CellarSafe requires no configuration of the gas monitors or alarm settings. Once the CellarSafe and repeater units are installed, power on the mains supply. The CellarSafe and repeater unit will become fully operational after a 40 second warm up period. The alarms and sounder will operate momentarily during this period and the operator display panel will test the display settings.

3.3.1 Carbon dioxide warning labels

The CellarSafe is provided with two carbon dioxide warning labels. It is important to fill in the contact and telephone details of the person responsible for the safety of the confined space and to affix the labels in clear, eye level positions, near the CellarSafe unit, at the entrance to the confined area being monitored, or near a repeater unit.

3.4 Self Test

A self test routine is provided to test the CellarSafe and any ancillary equipment attached via the alarm relays. To run the self test routine, press and hold the MODE button for 5 seconds. The self test can be initiated from the base unit or any of the repeater units.

The operator display panel will illuminate and test all display segments. The unit will enter alarm condition for a period of 10 seconds illuminating the alarm LEDS, activating the sounder and additional ancillary alarm equipment connected to the CellarSafe relays.

3.5 Alarm levels

The carbon dioxide and optional oxygen alarm thresholds are factory set at the following values:

	Low	High
Carbon dioxide, CO ₂	1.5%	3.0%
Oxygen, O ₂	19.5%	18.0%

3.6 In the event of a low or high alarm condition

If the low or high level alarm levels are exceeded the CellarSafe will activate the alarm relays, illuminate the alarm LEDs and activate the sounder on the base unit and all repeater units.

Carbon dioxide levels exceeded (CO₂): The sounder will beep once every second

Oxygen levels exceeded (O₂): The sounder will emit one long beep followed by one short beep once every second.

To silence the sounder, press and hold the MODE button for 5 seconds.

Note: CO₂ alarms are Subject to a 25 second time delay to prevent false alarms when exchanging gas cylinders. Oxygen alarms are instantaneous.

In the event of an alarm condition, DO NOT ENTER THE ROOM. Follow safety procedures as indicated on the carbon dioxide warning labels.

CellarSafe specification

Dimensions	210 x 150 x 75 mm
Weight	1.3 kg
Operating voltage	230 V, 50/60 Hz ac, 16 VA, double insulated 115v ac option available
Operating temperature	+3°C to +50°C (37°F to 122°F)
Warm up time	40 seconds
IP rating	IP 65
Display	Backlit LCD, 13.5 mm high digits
Instrument status indicators	2 gas level concentration alarm LEDs, red 1 fault alarm LED, yellow 1 power LED, green 1 sounder (>82 dB @ 10 cm)
Cable length	2 m
Fixing centres	4 at 192 mm x 72.5 mm
Humidity	0-99% RH non-condensing
Relay outputs	SPCO 10 A 240 V ac/30 V dc

Repeater unit specification

Dimensions	143 x 56 x 41 mm
Weight	120 g
IP rating	IP 54
Instrument status indicators	2 gas level concentration alarm LEDs, red 1 fault alarm LED, yellow 1 power LED, green 1 sounder (>82 dB @ 10 cm)
Cable length	9 m
Fixing centres	2 at 130 mm
Operating temperature	0°C to +50°C (32°F to 122°F)

5 TROUBLESHOOTING

Problem	Cause	Solution
Power LED off	Power not reaching processor	Check base unit is plugged in and electrical supply is switched on; check mains fuse
Power LED on continuously	Processor not working correctly	Call service engineer
Fault light on and sounder emitting rapid beeps	Sensor fault	Call service engineer
Operator display panel displaying an error	Base unit has been opened for servicing or wiring with battery option installed	Wait for 40 second warm up period. If error continues to be displayed call service engineer

6 CALIBRATION AND TESTING

6.1 Introduction

It is necessary to test this product at least every twelve months to ensure that the sensors and alarms are operational. Periodic testing may be performed simply by using exhaled breath to generate high CO₂ and low oxygen alarms on units with an oxygen sensor fitted. Exhaled breath contains approximately 4% CO₂ and 16% oxygen, and therefore contains sufficient levels of each gas to cause alarms on the CellarSafe monitor.

To test the alarms, blow into a plastic bag and hold the open end of the inflated bag tightly over the sensor(s) until the alarms activate. The CellarSafe display will show increased levels of CO₂ and decreased levels of oxygen (on units with an oxygen sensor fitted), and the alarms should activate after approximately 10-15 seconds for oxygen and 25 seconds for CO₂.

Alternatively, Crowcon can provide a calibration unit and calibration gas for checking and calibrating the sensors. The following sections describe the calibration procedure.

6.2 Calibration

These instructions provide guidelines for the service engineer to calibrate the CellarSafe using the CellarSafe calibration unit ("Figure 6.1"). The calibration unit is suitable for all models of the CellarSafe gas monitoring system. Calibration requires calibration gases and a pair of bayonet calibration adapters. In calibration mode all alarms and relays are inhibited.

Calibration gases

Zero calibration is performed using 100% nitrogen, and span calibration is performed using a gas mixture comprising 4.0% carbon dioxide and 20.75% oxygen with a balance of nitrogen.

Bayonet calibration adapters

A single pipe and bayonet fitting is provided for the CellarSafe carbon dioxide only monitoring system. The adapter for the CellarSafe carbon dioxide and oxygen monitoring system is supplied with a "T" junction and a single common supply pipe. The flow of the calibration gas should be set to 750 ml/min for the carbon dioxide only monitoring system and 1.5 l/min for the dual gas monitoring system.

Figure 6.1 CellarSafe Calibration Unit



6.3 Calibrating the CellarSafe gas monitor

Before connecting the Calibration Unit, ensure all power to the base unit is off. This includes disconnecting the battery if fitted. Failure to do so will risk damage to the unit.

Step 1:

Connecting the Calibration Unit

- 1 Switch off the mains supply to the CellarSafe base unit and any ancillary equipment connected to the alarm relays.
- 2 Remove the front cover and disconnect the battery if fitted.
- 3 Plug the lead from the calibration unit into the connector marked "PL2" on the printed circuit board.
- 4 Replace the front panel of the CellarSafe base unit. Do not over tighten the screws.
- 5 Reconnect the mains electrical supply.

Allow the instrument to stabilise for a few minutes. It is essential the CellarSafe reaches the appropriate thermal equilibrium for the gas monitors to work correctly and for the calibration to be performed without error.

NOTE: *In the unlikely event of a communication problem between the calibration unit and the base unit both LEDs will flicker rapidly to alert the user. If this occurs, the system may be reset by disconnecting the mains electrical supply for a few seconds (the battery should already be disconnected) and re-applying the mains power.*

Step 2:

Connecting the zero calibration gas

NOTE: *zero gas calibration MUST BE performed before the span gas calibration.*

- 1 Insert the bayonet of the adapter into the sensor inlets by gently twisting clockwise until they snap into place.
- 2 Connect the other end of the adapter to the zero calibration gas (100% nitrogen).
- 3 Adjust the gas flow to 750 ml/min for the carbon dioxide only monitoring system and 1.5 ml/min for the dual gas monitoring system.

Allow the system to stabilise for several minutes.

Step 3:**Calibrating zero gas**

- 1 Press the Scroll/Power button to turn on the calibration unit. The Zero Calibration or the Span Calibration LED will illuminate.

The base unit is now in calibration mode, the operator display panel will display alternately the internal instrument temperature and the gas reading.

- 2 If the Zero Calibration LED is not lit, press the Scroll/Power button.
- 3 Apply zero calibration gas as described in step 2 above, and wait for the gas reading in the operator display panel to stabilise.
- 4 Press the Select button to start the calibration. The Zero Calibration LED will flash to indicate that calibration is in progress.

After approximately one minute the Zero Calibration LED will stop flashing and zero calibration is complete.

NOTE: The carbon dioxide reading will briefly indicate an error on the operator display panel for a few seconds immediately after a calibration has been performed. This is normal.

To abort the calibration for any reason press the Scroll/Power button.

- 5 Proceed to step 4 to calibrate the span gas.

Step 4:**Calibrating span gas**

- 1 Connect the adapter to the span calibration gas.
- 2 Adjust the gas flow to 750 ml/min for the carbon dioxide only monitoring system and 1.5 l/min for the dual gas monitoring system.

Allow the system to stabilise for several minutes.

- 3 Press the Scroll/Power button to illuminate the Span calibration LED.

The base unit is now in calibration mode, the operator display panel will display alternately the internal instrument temperature and the gas reading.

- 4 Apply span calibration gas and wait for the gas reading in the operator display panel to stabilise.
- 5 Press the Select button to start the calibration. The Span Calibration LED will flash to indicate that calibration is in progress.

After approximately one minute the Span Calibration LED will stop flashing and span calibration is complete.

NOTE: The carbon dioxide reading will briefly indicate an error on the operator display panel for a few seconds immediately after a calibration has been performed. This is normal.

To abort the calibration for any reason press the Scroll/Power button.

- 6 Switch off the calibration unit by pressing the Scroll/Power button once, and then press and hold Scroll/Power button 3 seconds. Both LEDs will extinguish.

Step 5:**Calibration Unit disconnection**

- 1 Switch off the mains electrical supply to the base unit.
- 2 Disconnect the bayonet calibration adapter if not already done so.
- 3 Remove the four screws securing the front panel of the Base Unit.
- 4 Unplug the lead from the calibration unit from the printed circuit board.
- 5 If fitted, plug the battery into the connector marked "PL1" on the printed circuit board.
- 6 Refit the front panel of the base unit. Do not over tighten the screws.
- 7 Reconnect the mains electrical supply to the base unit and relays (if fitted).

The CellarSafe will perform the start up routine and stabilise after 2 minutes.



This product has been tested and found to comply with the European Directive 89/336/EEC, the low voltage directive EN61010 and the EMC requirements of EN50270.



Thank you for reading this data sheet.

For pricing or for further information, please contact us at our UK Office, using the details below.



UK Office
Keison Products,
P.O. Box 2124, Chelmsford, Essex, CM1 3UP, England.
Tel: +44 (0)330 088 0560
Fax: +44 (0)1245 808399
Email: sales@keison.co.uk

Please note - Product designs and specifications are subject to change without notice. The user is responsible for determining the suitability of this product.